

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100720\
 Data File : VN063965.D
 Acq On : 8 Oct 2020 1:41
 Operator : JC/MD
 Sample : PB132153ZHE#01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB132153ZHE#01

Quant Time: Oct 08 06:35:33 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	243521	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	451599	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	470766	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	187001	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	205758	51.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.28%	
35) Dibromofluoromethane	7.55	113	150747	48.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.22%	
50) Toluene-d8	10.06	98	582480	49.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.06%	
62) 4-Bromofluorobenzene	12.38	95	223077	49.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.92%	
Target Compounds						
16) Acetone	3.78	43	11838	9.559	ug/l	91
20) Methylene Chloride	4.50	84	10928	3.943	ug/l	96
43) Isopropyl Acetate	8.13	43	90977	12.134	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100720\
Data File : VN063965.D
Acq On : 8 Oct 2020 1:41
Operator : JC/MD
Sample : PB132153ZHE#01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB132153ZHE#01

Quant Time: Oct 08 06:35:33 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

