

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100720\
 Data File : VN063967.D
 Acq On : 8 Oct 2020 2:29
 Operator : JC/MD
 Sample : PB132153ZHE#03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB132153ZHE#03

Manual Integrations
 APPROVED

MMDadoda
 10/8/2020 10:09:50 AM

Quant Time: Oct 08 06:39:06 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	259259	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	476969	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	496038	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	201964	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	211252	49.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.64%	
35) Dibromofluoromethane	7.55	113	156501	47.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.58%	
50) Toluene-d8	10.06	98	615181	49.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.06%	
62) 4-Bromofluorobenzene	12.38	95	243291	51.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.18%	
Target Compounds						
16) Acetone	3.78	43	13060	9.905	ug/l	100
18) Methyl Acetate	4.29	43	3624	1.059	ug/l #	69
20) Methylene Chloride	4.51	84	10680m	3.628	ug/l	
43) Isopropyl Acetate	8.13	43	90707	11.455	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100720\
Data File : VN063967.D
Acq On : 8 Oct 2020 2:29
Operator : JC/MD
Sample : PB132153ZHE#03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB132153ZHE#03

Manual Integrations
APPROVED

MMDadoda
10/8/2020 10:09:50 AM

Quant Time: Oct 08 06:39:06 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

