

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100720\
 Data File : VN063970.D
 Acq On : 8 Oct 2020 3:41
 Operator : JC/MD
 Sample : L4224-05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OR-02-100620

Quant Time: Oct 08 06:46:21 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	245753	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	458054	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	469352	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	192409	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	207874	51.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.40%	
35) Dibromofluoromethane	7.55	113	150092	47.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.44%	
50) Toluene-d8	10.06	98	586332	48.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.32%	
62) 4-Bromofluorobenzene	12.38	95	223126	49.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.52%	
Target Compounds						
16) Acetone	3.78	43	15685	12.550	ug/l	96
18) Methyl Acetate	4.29	43	3948	1.217	ug/l #	69
20) Methylene Chloride	4.51	84	11529	4.117	ug/l	96
43) Isopropyl Acetate	8.13	43	16949	2.229	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100720\
Data File : VN063970.D
Acq On : 8 Oct 2020 3:41
Operator : JC/MD
Sample : L4224-05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OR-02-100620

Quant Time: Oct 08 06:46:21 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

