

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100720\
 Data File : VN063976.D
 Acq On : 8 Oct 2020 6:05
 Operator : JC/MD
 Sample : L4260-17
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FIELD-BLANK

Quant Time: Oct 08 07:03:12 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	234726	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	432121	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	441956	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	177351	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	204340	52.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.38%	
35) Dibromofluoromethane	7.55	113	151346	50.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.96%	
50) Toluene-d8	10.06	98	547821	48.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.40%	
62) 4-Bromofluorobenzene	12.38	95	212916	49.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.66%	
Target Compounds						
16) Acetone	3.78	43	10445	8.750	ug/l #	Qvalue 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100720\
Data File : VN063976.D
Acq On : 8 Oct 2020 6:05
Operator : JC/MD
Sample : L4260-17
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FIELD-BLANK

Quant Time: Oct 08 07:03:12 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

