

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100721\
 Data File : VN068841.D
 Acq On : 7 Oct 2021 15:07
 Operator : JC/MD
 Sample : M4057-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SHAFT-17B-SP

Manual Integrations
 APPROVED

MMDadoda
 10/8/2021 6:03:19 PM

Quant Time: Oct 08 05:58:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100621W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 06 14:41:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	479205	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	838090	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	806985	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	278756	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	335281	47.61	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	95.22%		
35) Dibromofluoromethane	8.021	113	260150	47.25	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	94.50%		
50) Toluene-d8	10.443	98	1081298	50.37	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	100.74%		
62) 4-Bromofluorobenzene	12.733	95	344953	43.23	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	86.46%		
Target Compounds						
3) Chloromethane	2.303	50	905	Below Cal		Qvalue 98
4) Vinyl Chloride	2.445	62	96	Below Cal	#	63
6) Chloroethane	3.057	64	460	Below Cal	#	44
16) Acetone	4.296	43	14057	Below Cal		96
17) Carbon Disulfide	4.545	76	26	Below Cal	#	74
18) Methyl Acetate	4.867	43	12944m	1.86	ug/l	
20) Methylene Chloride	5.106	84	23821	2.22	ug/l	89
25) 2-Butanone	7.345	43	6155	1.91	ug/l	95
31) Cyclohexane	8.085	56	12332	Below Cal	#	22
43) Isopropyl Acetate	8.566	43	63159	5.49	ug/l	# 74
96) 1,2,3-Trichlorobenzene	15.697	180	110	Below Cal	#	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100721\
Data File : VN068841.D
Acq On : 7 Oct 2021 15:07
Operator : JC/MD
Sample : M4057-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SHAFT-17B-SP

Manual Integrations
APPROVED

MMDadoda
10/8/2021 6:03:19 PM

Quant Time: Oct 08 05:58:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100621W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 06 14:41:09 2021
Response via : Initial Calibration

