

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100721\
 Data File : VN068844.D
 Acq On : 7 Oct 2021 16:22
 Operator : JC/MD
 Sample : M4111-05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 GROWS-2

Quant Time: Oct 08 05:59:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100621W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 06 14:41:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.091	168	435301	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	761230	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	695083	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	261785	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	310317	48.51	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.02%
35) Dibromofluoromethane	8.024	113	213584	42.71	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	85.42%
50) Toluene-d8	10.443	98	945713	48.50	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.00%
62) 4-Bromofluorobenzene	12.733	95	313679	43.27	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	86.54%
Target Compounds						
20) Methylene Chloride	5.101	84	29744	4.32	ug/l	Qvalue 92
25) 2-Butanone	7.343	43	9175	3.14	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100721\
Data File : VN068844.D
Acq On : 7 Oct 2021 16:22
Operator : JC/MD
Sample : M4111-05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
GROWS-2

Quant Time: Oct 08 05:59:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100621W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 06 14:41:09 2021
Response via : Initial Calibration

