

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100721\
 Data File : VN068846.D
 Acq On : 7 Oct 2021 17:13
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 08 06:00:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100621W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 06 14:41:09 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	135	0.00
2 T	Dichlorodifluoromethane	0.457	0.356	22.1	112	0.00
3 P	Chloromethane	0.470	0.382	18.7	117	0.00
4 C	Vinyl Chloride	0.885	0.793	10.4#	124	0.00
5 T	Bromomethane	0.792	0.684	13.6	121	-0.01
6 T	Chloroethane	0.750	0.661	11.9	120	-0.01
7 T	Trichlorofluoromethane	0.796	0.663	16.7	119	0.00
8 T	Diethyl Ether	0.263	0.226	14.1	120	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.445	0.372	16.4	116	0.00
10 T	Methyl Iodide	0.525	0.438	16.6	117	0.00
11 T	Tert butyl alcohol	0.087	0.083	4.6	131	0.00
12 CM	1,1-Dichloroethene	0.392	0.322	17.9#	116	0.00
13 T	Acrolein	0.047	0.034	27.7#	103	0.00
14 T	Allyl chloride	0.643	0.533	17.1	114	0.00
15 T	Acrylonitrile	0.229	0.210	8.3	122	0.00
16 T	Acetone	0.221	0.176	20.4	111	0.00
17 T	Carbon Disulfide	0.855	0.628	26.5#	115	0.00
18 T	Methyl Acetate	0.726	0.668	8.0	127	0.00
19 T	Methyl tert-butyl Ether	1.648	1.429	13.3	118	0.00
20 T	Methylene Chloride	0.519	0.423	18.5	115	0.00
21 T	trans-1,2-Dichloroethene	0.447	0.366	18.1	118	0.00
22 T	Diisopropyl ether	1.562	1.385	11.3	119	0.00
23 T	Vinyl Acetate	1.219	1.087	10.8	120	0.00
24 P	1,1-Dichloroethane	0.906	0.773	14.7	117	0.00
25 T	2-Butanone	0.335	0.301	10.1	120	0.00
26 T	2,2-Dichloropropane	0.867	0.685	21.0	107	0.00
27 T	cis-1,2-Dichloroethene	0.586	0.499	14.8	117	0.00
28 T	Bromochloromethane	0.492	0.426	13.4	116	0.00
29 T	Tetrahydrofuran	0.205	0.197	3.9	125	0.00
30 C	Chloroform	1.074	0.948	11.7#	120	0.00
31 T	Cyclohexane	0.764	0.586	23.3	119	0.00
32 T	1,1,1-Trichloroethane	0.956	0.843	11.8	119	0.00
33 S	1,2-Dichloroethane-d4	0.735	0.681	7.3	119	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	137	0.00
35 S	Dibromofluoromethane	0.328	0.295	10.1	118	0.00
36 T	1,1-Dichloropropene	0.395	0.334	15.4	116	0.00
37 T	Ethyl Acetate	0.377	0.353	6.4	124	0.00
38 T	Carbon Tetrachloride	0.476	0.409	14.1	120	0.00
39 T	Methylcyclohexane	0.443	0.362	18.3	117	0.00
40 TM	Benzene	1.235	1.054	14.7	119	0.00
41 T	Methacrylonitrile	0.189	0.164	13.2	118	0.00
42 TM	1,2-Dichloroethane	0.491	0.427	13.0	122	0.00
43 T	Isopropyl Acetate	0.686	0.648	5.5	127	0.00
44 TM	Trichloroethene	0.320	0.271	15.3	119	0.00
45 C	1,2-Dichloropropane	0.320	0.275	14.1#	117	0.00
46 T	Dibromomethane	0.249	0.218	12.4	120	0.00
47 T	Bromodichloromethane	0.500	0.444	11.2	120	0.00
48 T	Methyl methacrylate	0.293	0.269	8.2	124	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100721\
 Data File : VN068846.D
 Acq On : 7 Oct 2021 17:13
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 08 06:00:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100621W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 06 14:41:09 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.006	0.006	0.0	131	0.00
50 S	Toluene-d8	1.281	1.189	7.2	119	0.00
51 T	4-Methyl-2-Pentanone	0.437	0.415	5.0	128	0.00
52 CM	Toluene	0.831	0.724	12.9#	120	0.00
53 T	t-1,3-Dichloropropene	0.511	0.456	10.8	120	0.00
54 T	cis-1,3-Dichloropropene	0.526	0.458	12.9	120	0.00
55 T	1,1,2-Trichloroethane	0.346	0.315	9.0	122	0.00
56 T	Ethyl methacrylate	0.487	0.461	5.3	124	0.00
57 T	1,3-Dichloropropane	0.556	0.505	9.2	122	0.00
58 T	2-Chloroethyl Vinyl ether	0.122	0.143	-17.2	162#	0.00
59 T	2-Hexanone	0.315	0.300	4.8	129	0.00
60 T	Dibromochloromethane	0.388	0.355	8.5	122	0.00
61 T	1,2-Dibromoethane	0.352	0.319	9.4	122	0.00
62 S	4-Bromofluorobenzene	0.441	0.404	8.4	122	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	139	0.00
64 T	Tetrachloroethene	0.299	0.252	15.7	120	0.00
65 PM	Chlorobenzene	0.971	0.839	13.6	123	0.00
66 T	1,1,1,2-Tetrachloroethane	0.381	0.346	9.2	124	0.00
67 C	Ethyl Benzene	1.740	1.527	12.2#	122	0.00
68 T	m/p-Xylenes	0.661	0.577	12.7	122	0.00
69 T	o-Xylene	0.659	0.576	12.6	122	0.00
70 T	Styrene	1.097	0.981	10.6	123	0.00
71 P	Bromoform	0.280	0.260	7.1	127	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	142	0.00
73 T	Isopropylbenzene	3.950	3.443	12.8	122	0.00
74 T	N-amyl acetate	1.337	1.222	8.6	129	0.00
75 P	1,1,2,2-Tetrachloroethane	1.190	1.078	9.4	126	0.00
76 T	1,2,3-Trichloropropane	0.986	0.871	11.7	118	0.00
77 T	Bromobenzene	0.900	0.794	11.8	124	0.00
78 T	n-propylbenzene	4.493	3.961	11.8	124	0.00
79 T	2-Chlorotoluene	2.676	2.329	13.0	122	0.00
80 T	1,3,5-Trimethylbenzene	3.268	2.857	12.6	122	0.00
81 T	trans-1,4-Dichloro-2-butene	0.340	0.311	8.5	129	0.00
82 T	4-Chlorotoluene	2.655	2.274	14.4	121	0.00
83 T	tert-Butylbenzene	2.886	2.552	11.6	123	0.00
84 T	1,2,4-Trimethylbenzene	3.207	2.808	12.4	123	0.00
85 T	sec-Butylbenzene	4.109	3.627	11.7	125	0.00
86 T	p-Isopropyltoluene	3.326	2.910	12.5	124	0.00
87 T	1,3-Dichlorobenzene	1.688	1.432	15.2	122	0.00
88 T	1,4-Dichlorobenzene	1.666	1.380	17.2	122	0.00
89 T	n-Butylbenzene	2.806	2.434	13.3	126	0.00
90 T	Hexachloroethane	0.643	0.583	9.3	129	0.00
91 T	1,2-Dichlorobenzene	1.624	1.398	13.9	122	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.203	0.195	3.9	135	0.00
93 T	1,2,4-Trichlorobenzene	0.744	0.623	16.3	127	0.00
94 T	Hexachlorobutadiene	0.397	0.341	14.1	123	0.00
95 T	Naphthalene	1.985	1.713	13.7	131	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100721\
Data File : VN068846.D
Acq On : 7 Oct 2021 17:13
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Oct 08 06:00:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100621W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 06 14:41:09 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.721	0.619	14.1	130	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6