

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100724\
 Data File : VN084325.D
 Acq On : 07 Oct 2024 11:33
 Operator : JC\MD
 Sample : VN1007MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1007MBL01

Quant Time: Oct 08 03:56:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	154142	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	265135	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	224529	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	88778	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	111486	48.781	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.560%
35) Dibromofluoromethane	8.165	113	88228	50.476	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.960%
50) Toluene-d8	10.565	98	317446	49.363	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.720%
62) 4-Bromofluorobenzene	12.847	95	98197	41.914	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	83.820%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100724\
Data File : VN084325.D
Acq On : 07 Oct 2024 11:33
Operator : JC\MD
Sample : VN1007MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1007MBL01

Quant Time: Oct 08 03:56:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

