

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100724\
 Data File : VN084327.D
 Acq On : 07 Oct 2024 12:21
 Operator : JC\MD
 Sample : P4236-10
 Misc : 4.99g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP1-VOC

Quant Time: Oct 07 12:39:36 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	171198	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	292531	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	251901	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	106170	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	120949	47.649	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.300%
35) Dibromofluoromethane	8.159	113	95091	49.307	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.620%
50) Toluene-d8	10.565	98	352922	49.740	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.480%
62) 4-Bromofluorobenzene	12.847	95	111420	43.104	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	86.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100724\
Data File : VN084327.D
Acq On : 07 Oct 2024 12:21
Operator : JC\MD
Sample : P4236-10
Misc : 4.99g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP1-VOC

Quant Time: Oct 07 12:39:36 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

