

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100818\
 Data File : VN051730.D
 Acq On : 8 Oct 2018 12:55
 Operator : MD\SY
 Sample : J5165-03 10X
 Misc : 5.71g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 13 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 EP1-MW-F-092718(7-10)

Quant Time: Oct 09 14:05:01 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N100418W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 05 02:22:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	721280	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1044971	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	933122	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	607557	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	427853	44.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.80%	
35) Dibromofluoromethane	7.59	113	390474	48.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.66%	
50) Toluene-d8	10.09	98	1477324	49.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.04%	
62) 4-Bromofluorobenzene	12.40	95	587620	55.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.16%	
Target Compounds						
85) sec-Butylbenzene	13.17	105	648768	12.541	ug/l #	72
86) p-Isopropyltoluene	13.29	119	588229	12.826	ug/l	100
89) n-Butylbenzene	13.61	91	411383	11.345	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100818\
Data File : VN051730.D
Acq On : 8 Oct 2018 12:55
Operator : MD\SY
Sample : J5165-03 10X
Misc : 5.71g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 13 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
EP1-MW-F-092718(7-10)

Quant Time: Oct 09 14:05:01 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N100418W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 05 02:22:33 2018
Response via : Initial Calibration

