

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100819\
 Data File : VN058596.D
 Acq On : 8 Oct 2019 9:38
 Operator : JC/SP
 Sample : VSSTDICC0.1
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSSTDICC0.1

Quant Time: Oct 09 04:08:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N100819SIM-VINYL.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 09 04:00:06 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	3819726	5.00	ug/l	-0.02
4) 1,4-Difluorobenzene	8.56	114	5385210	5.00	ug/l	0.00
8) Chlorobenzene-d5	11.41	117	4617634	5.00	ug/l	0.00
9) 1,4-Dichlorobenzene-d4	13.35	152	1892496	5.00	ug/l	0.00

System Monitoring Compounds

3) 1,2-Dichloroethane-d4	0.00	65	0d	0.00	ug/l	
Spiked Amount	5.000	Range 81 - 118	Recovery	=	0.00%#	
5) Dibromofluoromethane	0.00	113	0d	0.00	ug/l	
Spiked Amount	5.000	Range 80 - 119	Recovery	=	0.00%#	
6) Toluene-d8	0.00	98	0d	0.00	ug/l	
Spiked Amount	5.000	Range 89 - 112	Recovery	=	0.00%#	
7) 4-Bromofluorobenzene	0.00	95	0d	0.00	ug/l	
Spiked Amount	5.000	Range 85 - 114	Recovery	=	0.00%#	

Target Compounds					Qvalue
2) Vinyl Chloride	2.16	62	33864	0.095 ug/l #	1

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100819\
Data File : VN058596.D
Acq On : 8 Oct 2019 9:38
Operator : JC/SP
Sample : VSSTDICC0.1
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSSTDICC0.1

Quant Time: Oct 09 04:08:01 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N100819SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Wed Oct 09 04:00:06 2019
Response via : Initial Calibration

