

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100819\
 Data File : VN058597.D
 Acq On : 8 Oct 2019 10:08
 Operator : JC/SP
 Sample : VSSTDICC0.5
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSSTDICC0.5

Quant Time: Oct 09 04:05:50 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N100819SIM-VINYL.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 09 04:00:06 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	3324403	5.00	ug/l	-0.02
4) 1,4-Difluorobenzene	8.56	114	4798563	5.00	ug/l	0.00
8) Chlorobenzene-d5	11.41	117	4138411	5.00	ug/l	0.00
9) 1,4-Dichlorobenzene-d4	13.35	152	1705589	5.00	ug/l	0.00

System Monitoring Compounds

3) 1,2-Dichloroethane-d4	8.00	65	200052	0.48	ug/l	0.00
Spiked Amount	5.000	Range	81 - 118	Recovery	=	9.60%#
5) Dibromofluoromethane	7.57	113	150403	0.48	ug/l	0.00
Spiked Amount	5.000	Range	80 - 119	Recovery	=	9.60%#
6) Toluene-d8	10.08	98	352460	0.54	ug/l	0.00
Spiked Amount	5.000	Range	89 - 112	Recovery	=	10.80%#
7) 4-Bromofluorobenzene	12.40	95	188202	1.04	ug/l	0.00
Spiked Amount	5.000	Range	85 - 114	Recovery	=	20.80%#

Target Compounds					Qvalue
2) Vinyl Chloride	2.16	62	179001	0.575	ug/l 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100819\
Data File : VN058597.D
Acq On : 8 Oct 2019 10:08
Operator : JC/SP
Sample : VSSTDICC0.5
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSSTDICC0.5

Quant Time: Oct 09 04:05:50 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N100819SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Wed Oct 09 04:00:06 2019
Response via : Initial Calibration

