

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100819\
 Data File : VN058598.D
 Acq On : 8 Oct 2019 10:30
 Operator : JC/SP
 Sample : VSSTDICC001
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSSTDICC001

Quant Time: Oct 09 04:05:56 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N100819SIM-VINYL.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 09 04:00:06 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	3177684	5.00	ug/l	0.00
4) 1,4-Difluorobenzene	8.56	114	4632033	5.00	ug/l	0.00
8) Chlorobenzene-d5	11.41	117	4001499	5.00	ug/l	0.00
9) 1,4-Dichlorobenzene-d4	13.35	152	1645605	5.00	ug/l	0.00
System Monitoring Compounds						
3) 1,2-Dichloroethane-d4	8.00	65	366452	0.92	ug/l	0.00
Spiked Amount 5.000	Range 81 - 118		Recovery	=	18.40%#	
5) Dibromofluoromethane	7.57	113	284313	0.94	ug/l	0.00
Spiked Amount 5.000	Range 80 - 119		Recovery	=	18.80%#	
6) Toluene-d8	10.08	98	646327	1.03	ug/l	0.00
Spiked Amount 5.000	Range 89 - 112		Recovery	=	20.60%#	
7) 4-Bromofluorobenzene	12.40	95	312994	1.79	ug/l	0.00
Spiked Amount 5.000	Range 85 - 114		Recovery	=	35.80%#	
Target Compounds						
2) Vinyl Chloride	2.16	62	334612	1.125	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100819\
Data File : VN058598.D
Acq On : 8 Oct 2019 10:30
Operator : JC/SP
Sample : VSSTDICC001
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSSTDICC001

Quant Time: Oct 09 04:05:56 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N100819SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Wed Oct 09 04:00:06 2019
Response via : Initial Calibration

