

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100819\
 Data File : VN058599.D
 Acq On : 8 Oct 2019 10:52
 Operator : JC/SP
 Sample : VSSTDICC002
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSSTDICC002

Quant Time: Oct 09 04:06:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N100819SIM-VINYL.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 09 04:00:06 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	3138976	5.00	ug/l	0.00
4) 1,4-Difluorobenzene	8.56	114	4589326	5.00	ug/l	0.00
8) Chlorobenzene-d5	11.41	117	3982456	5.00	ug/l	0.00
9) 1,4-Dichlorobenzene-d4	13.35	152	1624055	5.00	ug/l	0.00

System Monitoring Compounds

3) 1,2-Dichloroethane-d4	8.00	65	730870	1.86	ug/l	0.00
Spiked Amount	5.000	Range	81 - 118	Recovery	=	37.20%#
5) Dibromofluoromethane	7.57	113	568155	1.90	ug/l	0.00
Spiked Amount	5.000	Range	80 - 119	Recovery	=	38.00%#
6) Toluene-d8	10.08	98	1291020	2.08	ug/l	0.00
Spiked Amount	5.000	Range	89 - 112	Recovery	=	41.60%#
7) 4-Bromofluorobenzene	12.40	95	607490	3.51	ug/l	0.00
Spiked Amount	5.000	Range	85 - 114	Recovery	=	70.20%#

Target Compounds

	R.T.	QIon	Response	Conc	Units	Qvalue
2) Vinyl Chloride	2.16	62	627300	2.136	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100819\
Data File : VN058599.D
Acq On : 8 Oct 2019 10:52
Operator : JC/SP
Sample : VSSTDICC002
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSSTDICC002

Quant Time: Oct 09 04:06:01 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N100819SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Wed Oct 09 04:00:06 2019
Response via : Initial Calibration

