

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100819\
 Data File : VN058608.D
 Acq On : 8 Oct 2019 14:22
 Operator : JC/SP
 Sample : K5111-07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 LOD-MDL-WATER-01-QT4-2019

Quant Time: Oct 09 04:16:19 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N100819SIM-VINYL.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 09 04:07:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	2345991	5.00	ug/l	0.00
4) 1,4-Difluorobenzene	8.56	114	3475158	5.00	ug/l	0.00
8) Chlorobenzene-d5	11.41	117	3060735	5.00	ug/l	0.00
9) 1,4-Dichlorobenzene-d4	13.35	152	1185820	5.00	ug/l	0.00

System Monitoring Compounds						
3) 1,2-Dichloroethane-d4	8.00	65	1433585	5.11	ug/l	0.00
Spiked Amount	5.000	Range	81 - 118	Recovery	=	102.20%
5) Dibromofluoromethane	7.57	113	1099593	5.02	ug/l	0.00
Spiked Amount	5.000	Range	80 - 119	Recovery	=	100.40%
6) Toluene-d8	10.08	98	2475817	5.00	ug/l	0.00
Spiked Amount	5.000	Range	89 - 112	Recovery	=	100.00%
7) 4-Bromofluorobenzene	12.40	95	1161511	4.79	ug/l	0.00
Spiked Amount	5.000	Range	85 - 114	Recovery	=	95.80%

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Vinyl Chloride	2.16	62	11149	0.047	ug/l #	63

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100819\
Data File : VN058608.D
Acq On : 8 Oct 2019 14:22
Operator : JC/SP
Sample : K5111-07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
LOD-MDL-WATER-01-QT4-2019

Quant Time: Oct 09 04:16:19 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N100819SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Wed Oct 09 04:07:23 2019
Response via : Initial Calibration

