

Method Path : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\
 Method File : 82N100819SIM-VINYL.M
 Title : SW846 8260
 Last Update : Wed Oct 09 04:07:23 2019
 Response Via : Initial Calibration

Calibration Files

0.1 =VN058596.D 0.5 =VN058597.D 1 =VN058598.D
 5 =VN058600.D 8 =VN058601.D 2 =VN058599.D

	Compound	0.1	0.5	1	5	8	2	Avg	%RSD
1) I	Pentafluorobenzene	-----ISTD-----							
2) C	Vinyl Chloride	0.443	0.538	0.527	0.528	0.513	0.500	0.508	6.79#
3) S	1,2-Dichloroethan		0.602	0.577	0.618	0.608	0.582	0.597	2.94
4) I	1,4-Difluorobenzene	-----ISTD-----							
5) S	Dibromofluorometh		0.313	0.307	0.327	0.318	0.309	0.315	2.53
6) S	Toluene-d8		0.735	0.698	0.736	0.687	0.703	0.712	3.10
7) S	4-Bromofluorobenz		0.392	0.338	0.346	0.337	0.331	0.349	7.13
8) I	Chlorobenzene-d5	-----ISTD-----							
9) I	1,4-Dichlorobenzene-d	-----ISTD-----							

(#) = Out of Range