

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN100819\  
Data File : VN058602.D  
Acq On : 8 Oct 2019 11:58  
Operator : JC/SP  
Sample : VSSTDICV005  
Misc : 5.00mL/MSVOA N/WATER  
ALS Vial : 8 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
ICVVN100819

Quant Time: Oct 09 04:15:46 2019  
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_N\METHODS\82N100819SIM-VINYL.M  
Quant Title : SW846 8260  
QLast Update : Wed Oct 09 04:07:23 2019  
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
Max. RRF Dev : 20% Max. Rel. Area : 150%

|     | Compound               | Amount | Calc. | %Dev  | Area% | Dev(min) |
|-----|------------------------|--------|-------|-------|-------|----------|
| 1 I | Pentafluorobenzene     | 5.000  | 5.000 | 0.0   | 95    | 0.00     |
| 2 C | Vinyl Chloride         | 5.000  | 5.098 | -2.0# | 93    | 0.00     |
| 3 S | 1,2-Dichloroethane-d4  | 5.000  | 5.114 | -2.3  | 93    | 0.00     |
| 4 I | 1,4-Difluorobenzene    | 5.000  | 5.000 | 0.0   | 95    | 0.00     |
| 5 S | Dibromofluoromethane   | 5.000  | 5.054 | -1.1  | 92    | 0.00     |
| 6 S | Toluene-d8             | 5.000  | 5.122 | -2.4  | 94    | 0.00     |
| 7 S | 4-Bromofluorobenzene   | 5.000  | 4.951 | 1.0   | 95    | 0.00     |
| 8 I | Chlorobenzene-d5       | 5.000  | 5.000 | 0.0   | 95    | 0.00     |
| 9 I | 1,4-Dichlorobenzene-d4 | 5.000  | 5.000 | 0.0   | 95    | 0.00     |

(#) = Out of Range

SPCC's out = 0 CCC's out = 1