


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN100819\
Data File : VN058595.D
Acq On : 8 Oct 2019 9:08
Operator : JC/SP
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N100819SIM-VINYL.M
Title : SW846 8260
Last Update : Wed Oct 09 04:07:23 2019



AutoFind: Scans 3305, 3306, 3307; Background Corrected with Scan 3292

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.3	22141	PASS
75	95	30	60	51.9	56631	PASS
95	95	100	100	100.0	109120	PASS
96	95	5	9	6.6	7178	PASS
173	174	0.00	2	1.0	816	PASS
174	95	50	100	73.4	80069	PASS
175	174	5	9	7.8	6216	PASS
176	174	95	101	96.7	77442	PASS
177	176	5	9	7.3	5641	PASS