

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100820\
 Data File : VN064018.D
 Acq On : 9 Oct 2020 4:04
 Operator : JC/MD
 Sample : PB132189ZHE#07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB132189ZHE#07

Quant Time: Oct 09 07:06:04 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	240986	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	471792	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	484949	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	183034	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.99	65	211151	53.04	ug/l	0.00
Spiked Amount			Recovery	=	106.08%	
35) Dibromofluoromethane	7.55	113	153428	46.87	ug/l	0.00
Spiked Amount			Recovery	=	93.74%	
50) Toluene-d8	10.06	98	603226	48.61	ug/l	0.00
Spiked Amount			Recovery	=	97.22%	
62) 4-Bromofluorobenzene	12.38	95	230216	49.35	ug/l	0.00
Spiked Amount			Recovery	=	98.70%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
11) Tert butyl alcohol	4.73	59	128	0.291	ug/l #	71
16) Acetone	3.78	43	16851	13.750	ug/l	97
17) Carbon Disulfide	4.00	76	348	0.753	ug/l #	76
18) Methyl Acetate	4.29	43	3308	1.040	ug/l	92
20) Methylene Chloride	4.51	84	19005	6.855	ug/l	95
25) 2-Butanone	6.79	43	3998	2.301	ug/l #	83
29) Tetrahydrofuran	7.18	42	1077	0.924	ug/l #	38
31) Cyclohexane	7.62	56	5874	1.271	ug/l #	1
36) 1,1-Dichloropropene	7.62	75	21528	4.504	ug/l #	50
38) Carbon Tetrachloride	7.63	117	27318	5.097	ug/l #	15
43) Isopropyl Acetate	8.13	43	87795	11.209	ug/l	100
45) 1,2-Dichloropropane	9.23	63	1992	0.540	ug/l #	44
48) Methyl methacrylate	9.15	41	5092	1.316	ug/l #	12
49) 1,4-Dioxane	9.23	88	115	2.091	ug/l #	1
51) 4-Methyl-2-Pentanone	9.95	43	9191	2.000	ug/l	97
54) cis-1,3-Dichloropropene	9.87	75	34124	5.807	ug/l #	61
56) Ethyl methacrylate	10.51	69	2554	0.532	ug/l #	1
57) 1,3-Dichloropropane	10.74	76	6146	1.073	ug/l #	43
59) 2-Hexanone	10.74	43	106650	31.233	ug/l #	55
71) Bromoform	12.38	173	1335	0.458	ug/l #	1
74) N-amyl acetate	12.01	43	3229	2.713	ug/l #	51
76) 1,2,3-Trichloropropane	12.38	75	122812	30.438	ug/l #	100
80) 1,3,5-Trimethylbenzene	12.71	105	3833	0.342	ug/l	87
81) trans-1,4-Dichloro-2-buten	12.38	75	122812	87.667	ug/l #	6
95) Naphthalene	15.12	128	109	1.935	ug/l #	69


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100820\
Data File : VN064018.D
Acq On : 9 Oct 2020 4:04
Operator : JC/MD
Sample : PB132189ZHE#07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB132189ZHE#07

Quant Time: Oct 09 07:06:04 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.63 min Scan# 1831

Delta R.T. 0.00 min

Lab File: VN064018.D

Acq: 9 Oct 2020 4:04

Instrument :

MSVOA_N

ClientSampleId :


PB132189ZHE#07

Tot Ion:168 Resp: 240986

Ion Ratio Lower Upper

168 100

99 63.2 51.8 77.6



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN

7.63

100000

80000



60000

40000

20000

0

Time--> 7.50 7.60 7.70



#11

Tert butyl alcohol

Concen: 0.291 ug/l

RT: 4.73 min Scan# 931

Delta R.T. 0.01 min

Lab File: VN064018.D


Acq: 9 Oct 2020 4:04

Tgt Ion: 59 Resp: 128

Ion Ratio Lower Upper

59 100

57 0.0 8.7 13.1#



Abundance Ion 59.00 (58.70 to 59.70): VN

Ion 57.00 (56.70 to 57.70): VN

4.73

250

200


150


100

50

0

Time--> 4.72 4.73 4.74 4.75





#16

Acetone

Concen: 13.750 ug/l

RT: 3.78 min Scan# 635

Delta R.T. 0.01 min

Lab File: VN064018.D

Acq: 9 Oct 2020 4:04

Instrument :

MSVOA_N

ClientSampleId :


PB132189ZHE#07

Tot Ion: 43 Resp: 16851

Ion Ratio Lower Upper




43 100

58 26.2 22.3 33.5



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 58.00 (57.70 to 58.70): VN0



#17

Carbon Disulfide

Concen: 0.753 ug/l

RT: 4.00 min Scan# 704

Delta R.T. -0.01 min

Lab File: VN064018.D


Acq: 9 Oct 2020 4:04

Tgt Ion: 76 Resp: 348

Ion Ratio Lower Upper



76 100


78 0.0 7.0 10.4#



Abundance Ion 75.90 (75.60 to 76.60): VN0

Ion 77.90 (77.60 to 78.60): VN0









#18
Methyl Acetate
Concen: 1.040 ug/l
RT: 4.29 min Scan# 792
Delta R.T. 0.01 min
Lab File: VN064018.D
Acq: 9 Oct 2020 4:04




Instrument :
MSVOA_N
ClientSampleId :
PB132189ZHE#07


Tot Ion: 43 Resp: 3308
Ion Ratio Lower Upper
43 100
74 18.0 17.4 26.2



#20
Methylene Chloride
Concen: 6.855 ug/l
RT: 4.51 min Scan# 861
Delta R.T. 0.00 min
Lab File: VN064018.D
Acq: 9 Oct 2020 4:04

Tgt Ion: 84 Resp: 19005
Ion Ratio Lower Upper
84 100
49 130.9 110.4 165.6
51 39.5 33.3 49.9
86 61.1 50.7 76.1









#25
2-Butanone
Concen: 2.301 ug/l
RT: 6.79 min Scan# 1572
Delta R.T. 0.01 min
Lab File: VN064018.D
Acq: 9 Oct 2020 4:04




Instrument :
MSVOA_N
ClientSampleId :
PB132189ZHE#07


Tot Ion: 43 Resp: 3998
Ion Ratio Lower Upper
43 100
72 15.9 19.3 28.9#



#29
Tetrahydrofuran
Concen: 0.924 ug/l
RT: 7.18 min Scan# 1693
Delta R.T. 0.02 min
Lab File: VN064018.D
Acq: 9 Oct 2020 4:04

Tgt Ion: 42 Resp: 1077
Ion Ratio Lower Upper
42 100
72 0.0 32.8 49.2#
71 2.8 30.6 45.8#









#31
Cyclohexane
Concen: 1.271 ug/l
RT: 7.62 min Scan# 1830
Delta R.T. 0.01 min
Lab File: VN064018.D
Acq: 9 Oct 2020 4:04

Instrument :
MSVOA_N
ClientSampleId :
PB132189ZHE#07

Tot Ion: 56 Resp: 5874
Ion Ratio Lower Upper
56 100
69 161.9 25.4 38.2#
84 119.5 63.3 94.9#




Abundance Ion 56.00 (55.70 to 56.70): VN0
Ion 69.00 (68.70 to 69.70): VN0
Ion 84.00 (83.70 to 84.70): VN0






#33
1,2-Dichloroethane-d4
Concen: 53.036 ug/l
RT: 7.99 min Scan# 1943
Delta R.T. 0.00 min
Lab File: VN064018.D
Acq: 9 Oct 2020 4:04

Tgt Ion: 65 Resp: 211151
Ion Ratio Lower Upper
65 100
67 50.1 0.0 99.6



Abundance Ion 64.90 (64.60 to 65.60): VN0
Ion 66.90 (66.60 to 67.60): VN0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.55 min Scan# 2118

Delta R.T. 0.00 min

Lab File: VN064018.D

Acq: 9 Oct 2020 4:04

Instrument :

MSVOA_N

ClientSampleId :

PB132189ZHE#07


Tot Ion:114 Resp: 471792

Ion Ratio Lower Upper

114 100

63 22.0 0.0 45.0




88 16.5 0.0 34.0



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VN

Ion 87.90 (87.60 to 88.60): VN



#35

Dibromofluoromethane

Concen: 46.868 ug/l

RT: 7.55 min Scan# 1807

Delta R.T. 0.00 min

Lab File: VN064018.D

Acq: 9 Oct 2020 4:04


Tgt Ion:113 Resp: 153428

Ion Ratio Lower Upper

113 100

111 103.0 81.8 122.8



192 19.0 14.2 21.2




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V









#36
 1,1-Dichloropropene
 Concen: 4.504 ug/l
 RT: 7.62 min Scan# 1830
 Delta R.T. -0.13 min
 Lab File: VN064018.D
 Acq: 9 Oct 2020 4:04




Instrument :
 MSVOA_N
 ClientSampleId :
 PB132189ZHE#07


Tot Ion	Ratio	Lower	Upper
75	100		
110	8.5	16.9	50.6#
77	0.0	24.5	36.7#



#38
 Carbon Tetrachloride
 Concen: 5.097 ug/l
 RT: 7.63 min Scan# 1832
 Delta R.T. -0.10 min
 Lab File: VN064018.D
 Acq: 9 Oct 2020 4:04

Tgt Ion	Ratio	Lower	Upper
117	100		
119	5.6	80.1	120.1#
121	0.0	25.7	38.5#





#43

Isopropyl Acetate

Concen: 11.209 ug/l

RT: 8.13 min Scan# 1986

Delta R.T. 0.00 min

Lab File: VN064018.D

Acq: 9 Oct 2020 4:04

Instrument :

MSVOA_N

ClientSampleId :

PB132189ZHE#07


Tot Ion: 43 Resp: 87795

Ion Ratio Lower Upper

43 100

61 17.6 14.2 21.4




87 11.3 9.0 13.6



Abundance Ion 43.00 (42.70 to 43.70): VN064018.D (-1986) (-)

Ion 61.00 (60.70 to 61.70): VN064018.D (-1986) (-)

Ion 87.00 (86.70 to 87.70): VN064018.D (-1986) (-)



#45

1,2-Dichloropropane

Concen: 0.540 ug/l

RT: 9.23 min Scan# 2331

Delta R.T. 0.15 min

Lab File: VN064018.D


Acq: 9 Oct 2020 4:04

Tgt Ion: 63 Resp: 1992

Ion Ratio Lower Upper



63 100


65 0.0 24.7 37.1#



Abundance Ion 62.90 (62.60 to 63.60): VN064018.D (-2331) (-)

Ion 64.90 (64.60 to 65.60): VN064018.D (-2331) (-)









#48
Methyl methacrylate
Concen: 1.316 ug/l
RT: 9.15 min Scan# 2304
Delta R.T. -0.02 min
Lab File: VN064018.D
Acq: 9 Oct 2020 4:04

Instrument :
MSVOA_N
ClientSampleId :
PB132189ZHE#07

Tot Ion: 41 Resp: 5092
Ion Ratio Lower Upper
41 100
69 0.0 60.9 91.3#
39 0.0 57.5 86.3#




Abundance Ion 41.00 (40.70 to 41.70): VNO
Ion 69.00 (68.70 to 69.70): VNO
Ion 39.00 (38.70 to 39.70): VNO






#49
1,4-Dioxane
Concen: 2.091 ug/l
RT: 9.23 min Scan# 2331
Delta R.T. 0.07 min
Lab File: VN064018.D
Acq: 9 Oct 2020 4:04

Tgt Ion: 88 Resp: 115
Ion Ratio Lower Upper
88 100
43 1458155.7 33.1 49.7#
58 6570.4 58.9 88.3#



Abundance Ion 88.00 (87.70 to 88.70): VNO
Ion 43.00 (42.70 to 43.70): VNO
Ion 58.00 (57.70 to 58.70): VNO





#50

Toluene-d8

Concen: 48.608 ug/l

RT: 10.06 min Scan# 2588

Delta R.T. 0.00 min

Lab File: VN064018.D

Acq: 9 Oct 2020 4:04

Instrument :

MSVOA_N

ClientSampleId :





PB132189ZHE#07

Tot Ion: 98 Resp: 603226

Ion Ratio Lower Upper

98 100

100 63.4 49.8 74.6



#51

4-Methyl-2-Pentanone

Concen: 2.000 ug/l

RT: 9.95 min Scan# 2554

Delta R.T. 0.00 min

Lab File: VN064018.D




Acq: 9 Oct 2020 4:04


Tgt Ion: 43 Resp: 9191

Ion Ratio Lower Upper

43 100

58 33.5 28.4 42.6





#54

cis-1,3-Dichloropropene

Concen: 5.807 ug/l

RT: 9.87 min Scan# 2530

Delta R.T. 0.07 min

Lab File: VN064018.D

Acq: 9 Oct 2020 4:04

Instrument :

MSVOA_N

ClientSampleId :

PB132189ZHE#07





Tot Ion: 75 Resp: 34124

Ion Ratio Lower Upper

75 100

77 0.0 24.6 36.8#

39 38.6 49.7 74.5#



#56

Ethyl methacrylate

Concen: 0.532 ug/l

RT: 10.51 min Scan# 2729

Delta R.T. 0.12 min

Lab File: VN064018.D

Acq: 9 Oct 2020 4:04




Tgt Ion: 69 Resp: 2554


Ion Ratio Lower Upper

69 100

41 3531.7 61.1 91.7#

39 1530.8 42.2 63.4#









#57
 1,3-Dichloropropane
 Concen: 1.073 ug/l
 RT: 10.74 min Scan# 2798
 Delta R.T. 0.06 min
 Lab File: VN064018.D
 Acq: 9 Oct 2020 4:04




Instrument :
 MSVOA_N
 ClientSampleId :
 PB132189ZHE#07


Tot Ion: 76 Resp: 6146
 Ion Ratio Lower Upper
 76 100
 78 0.0 25.2 37.8#



#59
 2-Hexanone
 Concen: 31.233 ug/l
 RT: 10.74 min Scan# 2798
 Delta R.T. 0.02 min
 Lab File: VN064018.D
 Acq: 9 Oct 2020 4:04

Tgt Ion: 43 Resp: 106650
 Ion Ratio Lower Upper
 43 100
 58 18.6 24.6 73.8#





#62

4-Bromofluorobenzene

Concen: 49.349 ug/l

RT: 12.38 min Scan# 3308

Delta R.T. 0.00 min

Lab File: VN064018.D

Acq: 9 Oct 2020 4:04

Instrument :

MSVOA_N

ClientSampleId :

PB132189ZHE#07


Tot Ion: 95 Resp: 230216

Ion Ratio Lower Upper

95 100

174 77.1 0.0 150.6




176 72.2 0.0 145.4



Abundance Ion 94.90 (94.60 to 95.60): VN063741.D (-3293) (-)

Ion 173.80 (173.50 to 174.50): VN063741.D (-3293) (-)

Ion 175.80 (175.50 to 176.50): VN063741.D (-3293) (-)



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.38 min Scan# 2998

Delta R.T. 0.00 min

Lab File: VN064018.D

Acq: 9 Oct 2020 4:04


Tgt Ion: 117 Resp: 484949

Ion Ratio Lower Upper

117 100

82 59.2 47.4 71.2



119 32.3 26.4 39.6




Abundance Ion 116.90 (116.60 to 117.60): VN063741.D (-2983) (-)

Ion 82.00 (81.70 to 82.70): VN063741.D (-2983) (-)

Ion 118.90 (118.60 to 119.60): VN063741.D (-2983) (-)





#71

Bromoform

Concen: 0.458 ug/l

RT: 12.38 min Scan# 3310

Delta R.T. 0.28 min

Lab File: VN064018.D

Acq: 9 Oct 2020 4:04

Instrument :

MSVOA_N

ClientSampleId :

PB132189ZHE#07


Tot Ion:173 Resp: 1335

Ion Ratio Lower Upper




173 100

175 816.0 23.5 70.5#

254 0.0 0.0 0.0



Abundance Ion 172.70 (172.40 to 173.40): V
Ion 174.70 (174.40 to 175.40): V
Ion 254.40 (254.10 to 255.10): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.32 min Scan# 3601

Delta R.T. 0.00 min

Lab File: VN064018.D

Acq: 9 Oct 2020 4:04


Tgt Ion:152 Resp: 183034

Ion Ratio Lower Upper



152 100


115 62.6 31.3 93.8

150 159.7 0.0 341.8



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V





#74

N-amvl acetate

Concen: 2.713 ug/l

RT: 12.01 min Scan# 3193

Delta R.T. -0.04 min

Lab File: VN064018.D

Acq: 9 Oct 2020 4:04

Instrument :

MSVOA_N

ClientSampleId :

PB132189ZHE#07

Tot Ion: 43 Resp: 3229


Ion Ratio Lower Upper

43 100

70 0.0 29.8 44.8#

55 7.9 17.9 26.9#

61 0.0 17.0 25.4#






Abundance Ion 43.00 (42.70 to 43.70): VNO

Ion 70.00 (69.70 to 70.70): VNO

Ion 55.00 (54.70 to 55.70): VNO

Ion 61.00 (60.70 to 61.70): VNO



#76

1,2,3-Trichloropropane

Concen: 30.438 ug/l

RT: 12.38 min Scan# 3308

Delta R.T. -0.15 min

Lab File: VN064018.D


Acq: 9 Oct 2020 4:04

Tgt Ion: 75 Resp: 122812

Ion Ratio Lower Upper



75 100


77 1.2 0.0 0.0#



Abundance Ion 74.90 (74.60 to 75.60): VNO

Ion 77.00 (76.70 to 77.70): VNO









#80
 1,3,5-Trimethylbenzene
 Concen: 0.342 ug/l
 RT: 12.71 min Scan# 3412
 Delta R.T. 0.00 min
 Lab File: VN064018.D
 Acq: 9 Oct 2020 4:04




Instrument :
 MSVOA_N
 ClientSampleId :
 PB132189ZHE#07


Tot Ion:105 Resp: 3833
 Ion Ratio Lower Upper
 105 100
 120 55.0 23.3 69.9



#81
 trans-1,4-Dichloro-2-butene
 Concen: 87.667 ug/l
 RT: 12.38 min Scan# 3308
 Delta R.T. 0.10 min
 Lab File: VN064018.D
 Acq: 9 Oct 2020 4:04

Tgt Ion: 75 Resp: 122812
 Ion Ratio Lower Upper
 75 100
 53 0.1 88.2 132.2#
 89 0.0 36.0 54.0#






#95
Naphthalene
Concen: 1.935 ug/l
RT: 15.12 min Scan# 4161
Delta R.T. 0.02 min
Lab File: VN064018.D
Acq: 9 Oct 2020 4:04

Instrument :
MSVOA_N
ClientSampleId :
PB132189ZHE#07

Tot Ion	Ratio	Lower	Upper
128	100		
127	0.0	10.5	15.7#
129	0.0	8.7	13.1#



Abundance

Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V

