

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100820\
 Data File : VN064012.D
 Acq On : 9 Oct 2020 1:40
 Operator : JC/MD
 Sample : PB132189ZHE#01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB132189ZHE#01

Quant Time: Oct 09 07:18:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	242727	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	469886	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	487930	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	186608	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	203000	50.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.24%	
35) Dibromofluoromethane	7.55	113	153011	46.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.86%	
50) Toluene-d8	10.06	98	595476	48.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.36%	
62) 4-Bromofluorobenzene	12.38	95	228949	49.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.56%	
Target Compounds						
16) Acetone	3.79	43	13810	11.188	ug/l	98
18) Methyl Acetate	4.29	43	4549	1.419	ug/l #	86
20) Methylene Chloride	4.51	84	8950	3.257	ug/l	98
43) Isopropyl Acetate	8.13	43	94237	12.080	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100820\
Data File : VN064012.D
Acq On : 9 Oct 2020 1:40
Operator : JC/MD
Sample : PB132189ZHE#01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB132189ZHE#01

Quant Time: Oct 09 07:18:14 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

