

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100820\
 Data File : VN064013.D
 Acq On : 9 Oct 2020 2:04
 Operator : JC/MD
 Sample : PB132189ZHE#02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB132189ZHE#02

Quant Time: Oct 09 07:20:13 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	246640	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	461481	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	480037	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	182798	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	205703	50.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.96%	
35) Dibromofluoromethane	7.55	113	149572	46.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.42%	
50) Toluene-d8	10.06	98	591173	48.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.40%	
62) 4-Bromofluorobenzene	12.38	95	226421	49.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.24%	
Target Compounds						
16) Acetone	3.78	43	13665	10.894	ug/l	98
18) Methyl Acetate	4.29	43	4167	1.280	ug/l #	54
20) Methylene Chloride	4.50	84	8643	3.100	ug/l #	90
43) Isopropyl Acetate	8.13	43	94926	12.390	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100820\
Data File : VN064013.D
Acq On : 9 Oct 2020 2:04
Operator : JC/MD
Sample : PB132189ZHE#02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB132189ZHE#02

Quant Time: Oct 09 07:20:13 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

