

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100820\
 Data File : VN064021.D
 Acq On : 9 Oct 2020 5:16
 Operator : JC/MD
 Sample : L4291-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PAV-B4-100620-19-20

Manual Integrations
 APPROVED

MMDadoda
 10/9/2020 4:53:23 PM

Quant Time: Oct 09 09:44:30 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	248057	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	473442	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	493697	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	189368	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	207114	50.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.08%	
35) Dibromofluoromethane	7.55	113	157915	48.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.14%	
50) Toluene-d8	10.06	98	609941	48.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.96%	
62) 4-Bromofluorobenzene	12.38	95	231655	49.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.96%	
Target Compounds						
16) Acetone	3.78	43	15720	12.461	ug/l	96
20) Methylene Chloride	4.51	84	10375m	3.682	ug/l	
43) Isopropyl Acetate	8.13	43	89829	11.428	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100820\
Data File : VN064021.D
Acq On : 9 Oct 2020 5:16
Operator : JC/MD
Sample : L4291-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PAV-B4-100620-19-20

Manual Integrations
APPROVED

MMDadoda
10/9/2020 4:53:23 PM

Quant Time: Oct 09 09:44:30 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

