

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100820\
 Data File : VN064027.D
 Acq On : 9 Oct 2020 7:40
 Operator : JC/MD
 Sample : L4311-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PAV-B10-100720-16-17

Quant Time: Oct 09 12:26:47 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	253777	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	487085	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	497632	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	189072	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	211006	50.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.66%	
35) Dibromofluoromethane	7.55	113	156536	46.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.64%	
50) Toluene-d8	10.06	98	614025	47.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.84%	
62) 4-Bromofluorobenzene	12.38	95	229902	47.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.46%	
Target Compounds						
16) Acetone	3.79	43	15607	12.093	ug/l	96
20) Methylene Chloride	4.51	84	10604	3.678	ug/l	93
43) Isopropyl Acetate	8.13	43	86330	10.676	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100820\
Data File : VN064027.D
Acq On : 9 Oct 2020 7:40
Operator : JC/MD
Sample : L4311-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 54 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PAV-B10-100720-16-17

Quant Time: Oct 09 12:26:47 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

