

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100820\
 Data File : VN064028.D
 Acq On : 9 Oct 2020 8:04
 Operator : JC/MD
 Sample : L4299-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 86-96-COMP

Quant Time: Oct 09 12:29:11 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	236085	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	444848	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	444592	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	175478	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	197379	50.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.22%	
35) Dibromofluoromethane	7.55	113	150308	48.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.40%	
50) Toluene-d8	10.06	98	554785	47.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.82%	
62) 4-Bromofluorobenzene	12.38	95	211823	48.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.32%	
Target Compounds						
11) Tert butyl alcohol	4.73	59	88165	204.747	ug/l #	Qvalue 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100820\
Data File : VN064028.D
Acq On : 9 Oct 2020 8:04
Operator : JC/MD
Sample : L4299-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 55 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
86-96-COMP

Quant Time: Oct 09 12:29:11 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

