

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN100820\
 Data File : VN064004.D
 Acq On : 8 Oct 2020 20:29
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 09 06:52:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	103	0.00
2 T	Dichlorodifluoromethane	0.528	0.451	14.6	94	0.00
3 P	Chloromethane	0.619	0.609	1.6	112	0.00
4 C	Vinyl Chloride	0.621	0.581	6.4#	105	0.00
5 T	Bromomethane	0.391	0.312	20.2	90	0.00
6 T	Chloroethane	0.371	0.369	0.5	108	0.00
7 T	Trichlorofluoromethane	0.986	0.935	5.2	104	0.00
8 T	Diethyl Ether	0.333	0.336	-0.9	105	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.527	0.488	7.4	101	0.00
10 T	Methyl Iodide	0.612	0.569	7.0	94	0.00
11 T	Tert butyl alcohol	0.091	0.099	-8.8	113	0.00
12 CM	1,1-Dichloroethene	0.503	0.481	4.4#	106	0.00
13 T	Acrolein	0.056	0.071	-26.8#	128	0.00
14 T	Allyl chloride	0.935	0.902	3.5	103	0.00
15 T	Acrylonitrile	0.261	0.292	-11.9	117	0.00
16 T	Acetone	0.254	0.264	-3.9	111	0.00
17 T	Carbon Disulfide	1.465	1.248	14.8	99	0.00
18 T	Methyl Acetate	0.660	0.762	-15.5	128	0.00
19 T	Methyl tert-butyl Ether	1.829	1.904	-4.1	110	0.00
20 T	Methylene Chloride	0.630	0.584	7.3	106	0.00
21 T	trans-1,2-Dichloroethene	0.564	0.532	5.7	105	0.00
22 T	Diisopropyl ether	1.925	2.160	-12.2	118	0.00
23 T	Vinyl Acetate	1.594	1.742	-9.3	111	0.00
24 P	1,1-Dichloroethane	1.100	1.106	-0.5	107	0.00
25 T	2-Butanone	0.361	0.409	-13.3	116	0.00
26 T	2,2-Dichloropropane	1.008	0.843	16.4	90	0.00
27 T	cis-1,2-Dichloroethene	0.649	0.641	1.2	107	0.00
28 T	Bromochloromethane	0.544	0.578	-6.2	110	0.00
29 T	Tetrahydrofuran	0.242	0.278	-14.9	120	0.00
30 C	Chloroform	1.147	1.144	0.3#	107	0.00
31 T	Cyclohexane	0.959	0.892	7.0	104	0.00
32 T	1,1,1-Trichloroethane	1.035	1.023	1.2	103	0.00
33 S	1,2-Dichloroethane-d4	0.826	0.763	7.6	99	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	105	0.00
35 S	Dibromofluoromethane	0.347	0.324	6.6	101	0.00
36 T	1,1-Dichloropropene	0.507	0.479	5.5	104	0.00
37 T	Ethyl Acetate	0.503	0.534	-6.2	119	0.00
38 T	Carbon Tetrachloride	0.568	0.517	9.0	101	0.00
39 T	Methylcyclohexane	0.505	0.459	9.1	97	0.00
40 TM	Benzene	1.447	1.427	1.4	108	0.00
41 T	Methacrylonitrile	0.224	0.241	-7.6	106	0.00
42 TM	1,2-Dichloroethane	0.618	0.579	6.3	102	0.00
43 T	Isopropyl Acetate	0.830	0.889	-7.1	111	0.00
44 TM	Trichloroethene	0.379	0.355	6.3	104	0.00
45 C	1,2-Dichloropropane	0.391	0.403	-3.1#	111	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN100820\
 Data File : VN064004.D
 Acq On : 8 Oct 2020 20:29
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 09 06:52:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.270	0.255	5.6	104	0.00
47 T	Bromodichloromethane	0.556	0.549	1.3	105	0.00
48 T	Methyl methacrylate	0.410	0.452	-10.2	117	0.00
49 T	1,4-Dioxane	0.006	0.006	0.0	115	0.00
50 S	Toluene-d8	1.315	1.219	7.3	102	0.00
51 T	4-Methyl-2-Pentanone	0.487	0.546	-12.1	114	0.00
52 CM	Toluene	0.895	0.900	-0.6#	107	0.00
53 T	t-1,3-Dichloropropene	0.596	0.587	1.5	103	0.00
54 T	cis-1,3-Dichloropropene	0.623	0.620	0.5	104	0.00
55 T	1,1,2-Trichloroethane	0.359	0.363	-1.1	107	0.00
56 T	Ethyl methacrylate	0.509	0.563	-10.6	113	0.00
57 T	1,3-Dichloropropane	0.607	0.618	-1.8	107	0.00
58 T	2-Chloroethyl Vinyl ether	0.272	0.282	-3.7	107	0.00
59 T	2-Hexanone	0.362	0.390	-7.7	112	0.00
60 T	Dibromochloromethane	0.418	0.413	1.2	107	0.00
61 T	1,2-Dibromoethane	0.367	0.370	-0.8	105	0.00
62 S	4-Bromofluorobenzene	0.494	0.458	7.3	101	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	105	0.00
64 T	Tetrachloroethene	0.387	0.344	11.1	103	0.00
65 PM	Chlorobenzene	1.026	1.003	2.2	108	0.00
66 T	1,1,1,2-Tetrachloroethane	0.393	0.389	1.0	106	0.00
67 C	Ethyl Benzene	1.866	1.836	1.6#	104	0.00
68 T	m/p-Xylenes	0.696	0.672	3.4	103	0.00
69 T	o-Xylene	0.650	0.646	0.6	105	0.00
70 T	Styrene	1.104	1.141	-3.4	106	0.00
71 P	Bromoform	0.301	0.301	0.0	105	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
73 T	Isopropylbenzene	3.647	3.620	0.7	101	0.00
74 T	N-amyl acetate	1.509	2.031	-34.6#	129	0.00
75 P	1,1,2,2-Tetrachloroethane	1.115	1.145	-2.7	107	0.00
76 T	1,2,3-Trichloropropane	1.102	1.046	5.1	106	0.00
77 T	Bromobenzene	0.895	0.916	-2.3	107	0.00
78 T	n-propylbenzene	4.244	4.124	2.8	101	0.00
79 T	2-Chlorotoluene	2.610	2.556	2.1	104	0.00
80 T	1,3,5-Trimethylbenzene	3.060	3.059	0.0	102	0.00
81 T	trans-1,4-Dichloro-2-butene	0.383	0.392	-2.3	102	0.00
82 T	4-Chlorotoluene	2.710	2.657	2.0	101	0.00
83 T	tert-Butylbenzene	2.503	2.481	0.9	102	0.00
84 T	1,2,4-Trimethylbenzene	3.055	3.160	-3.4	103	0.00
85 T	sec-Butylbenzene	3.287	3.210	2.3	100	0.00
86 T	p-Isopropyltoluene	2.908	2.886	0.8	99	0.00
87 T	1,3-Dichlorobenzene	1.619	1.573	2.8	101	0.00
88 T	1,4-Dichlorobenzene	1.686	1.590	5.7	103	0.00
89 T	n-Butylbenzene	2.684	2.464	8.2	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN100820\
Data File : VN064004.D
Acq On : 8 Oct 2020 20:29
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Oct 09 06:52:56 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.564	0.539	4.4	97	0.00
91 T	1,2-Dichlorobenzene	1.604	1.571	2.1	104	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.262	0.249	5.0	100	0.00
93 T	1,2,4-Trichlorobenzene	0.881	0.754	14.4	92	0.00
94 T	Hexachlorobutadiene	0.511	0.424	17.0	100	0.00
95 T	Naphthalene	2.699	2.716	-0.6	107	0.00
96 T	1,2,3-Trichlorobenzene	0.883	0.773	12.5	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6