

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100820\
 Data File : VN063985.D
 Acq On : 8 Oct 2020 12:55
 Operator : JC/MD
 Sample : L4255-20
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB01-20201001

Quant Time: Oct 09 04:54:11 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	234505	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	411029	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	420703	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	160173	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	190429	49.15	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.30%	
35) Dibromofluoromethane	7.55	113	141318	49.55	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.10%	
50) Toluene-d8	10.06	98	524436	48.51	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.02%	
62) 4-Bromofluorobenzene	12.38	95	195704	48.15	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.30%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100820\
Data File : VN063985.D
Acq On : 8 Oct 2020 12:55
Operator : JC/MD
Sample : L4255-20
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB01-20201001

Quant Time: Oct 09 04:54:11 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.63 min Scan# 1831

Delta R.T. 0.00 min

Lab File: VN063985.D

Acq: 8 Oct 2020 12:55

Instrument :

MSVOA_N

ClientSampleId :


TB01-20201001

Tot Ion:168 Resp: 234505

Ion Ratio Lower Upper

168 100

99 62.6 51.8 77.6



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO

7.63

100000

80000



60000

40000

20000

0

Time-->



#33

1,2-Dichloroethane-d4

Concen: 49.153 ug/l

RT: 7.99 min Scan# 1943

Delta R.T. 0.00 min

Lab File: VN063985.D


Acq: 8 Oct 2020 12:55

Tgt Ion: 65 Resp: 190429

Ion Ratio Lower Upper

65 100

67 49.9 0.0 99.6



Abundance Ion 64.90 (64.60 to 65.60): VNO

Ion 66.90 (66.60 to 67.60): VNO

7.99

80000


60000


40000

20000

0

Time-->





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.55 min Scan# 2118

Delta R.T. 0.00 min

Lab File: VN063985.D

Acq: 8 Oct 2020 12:55

Instrument :

MSVOA_N

ClientSampleId :

TB01-20201001


Tot Ion:114 Resp: 411029

Ion Ratio Lower Upper

114 100

63 23.0 0.0 45.0




88 15.9 0.0 34.0



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VN

Ion 87.90 (87.60 to 88.60): VN



#35

Dibromofluoromethane

Concen: 49.550 ug/l

RT: 7.55 min Scan# 1807

Delta R.T. 0.00 min

Lab File: VN063985.D

Acq: 8 Oct 2020 12:55


Tgt Ion:113 Resp: 141318

Ion Ratio Lower Upper

113 100

111 102.3 81.8 122.8



192 18.8 14.2 21.2




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V









#50
Toluene-d8
Concen: 48.506 ug/l
RT: 10.06 min Scan# 2587
Delta R.T. 0.00 min
Lab File: VN063985.D
Acq: 8 Oct 2020 12:55




Instrument :
MSVOA_N
ClientSampleId :
TB01-20201001


Tot Ion: 98 Resp: 524436
Ion Ratio Lower Upper
98 100
100 63.2 49.8 74.6



#62
4-Bromofluorobenzene
Concen: 48.153 ug/l
RT: 12.38 min Scan# 3308
Delta R.T. 0.00 min
Lab File: VN063985.D
Acq: 8 Oct 2020 12:55

Tgt Ion: 95 Resp: 195704
Ion Ratio Lower Upper
95 100
174 78.0 0.0 150.6
176 74.7 0.0 145.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.38 min Scan# 2998
Delta R.T. 0.00 min
Lab File: VN063985.D
Acq: 8 Oct 2020 12:55

Instrument :
MSVOA_N
ClientSampleId :
TB01-20201001

Tot Ion:117 Resp: 420703
Ion Ratio Lower Upper
117 100
82 58.6 47.4 71.2
119 31.3 26.4 39.6



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.32 min Scan# 3601
Delta R.T. 0.00 min
Lab File: VN063985.D
Acq: 8 Oct 2020 12:55

Tgt Ion:152 Resp: 160173
Ion Ratio Lower Upper
152 100
115 62.8 31.3 93.8
150 158.0 0.0 341.8

