

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100820\
 Data File : VN063996.D
 Acq On : 8 Oct 2020 17:18
 Operator : JC/MD
 Sample : L4255-07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SA-PZ148-20200930

Quant Time: Oct 09 06:32:43 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	251895	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	458759	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	476460	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	189891	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	209363	50.31	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.62%	
35) Dibromofluoromethane	7.55	113	159963	50.25	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.50%	
50) Toluene-d8	10.06	98	595074	49.31	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.62%	
62) 4-Bromofluorobenzene	12.38	95	224461	49.48	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.96%	


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100820\
Data File : VN063996.D
Acq On : 8 Oct 2020 17:18
Operator : JC/MD
Sample : L4255-07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SA-PZ148-20200930

Quant Time: Oct 09 06:32:43 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.63 min Scan# 1831

Delta R.T. 0.00 min

Lab File: VN063996.D

Acq: 8 Oct 2020 17:18

Instrument :

MSVOA_N

ClientSampleId :


SA-PZ148-20200930

Tot Ion:168 Resp: 251895

Ion Ratio Lower Upper

168 100

99 61.8 51.8 77.6



Abundance Ion 167.90 (167.60 to 168.60): VN063996.D (-1738) (-)

Ion 98.90 (98.60 to 99.60): VN063996.D (-1738) (-)

7.63

120000

100000

80000



60000

40000

20000

0

Time--> 7.50 7.60 7.70



#33

1,2-Dichloroethane-d4

Concen: 50.310 ug/l

RT: 7.99 min Scan# 1943

Delta R.T. 0.00 min

Lab File: VN063996.D


Acq: 8 Oct 2020 17:18

Tgt Ion: 65 Resp: 209363

Ion Ratio Lower Upper

65 100

67 50.6 0.0 99.6



Abundance Ion 64.90 (64.60 to 65.60): VN063996.D (-1850) (-)

Ion 66.90 (66.60 to 67.60): VN063996.D (-1850) (-)

7.99

100000

80000


60000


40000

20000

0

Time--> 7.90 7.95 8.00 8.05 8.10









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.55 min Scan# 2118
 Delta R.T. 0.00 min
 Lab File: VN063996.D
 Acq: 8 Oct 2020 17:18

Instrument :
 MSVOA_N
 ClientSampleId :
 SA-PZ148-20200930

Tot Ion:114 Resp: 458759
 Ion Ratio Lower Upper
 114 100
 63 23.2 0.0 45.0
 88 16.3 0.0 34.0




Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): V
 Ion 87.90 (87.60 to 88.60): V






#35
 Dibromofluoromethane
 Concen: 50.252 ug/l
 RT: 7.55 min Scan# 1807
 Delta R.T. 0.00 min
 Lab File: VN063996.D
 Acq: 8 Oct 2020 17:18

Tgt Ion:113 Resp: 159963
 Ion Ratio Lower Upper
 113 100
 111 102.5 81.8 122.8
 192 19.3 14.2 21.2



Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 49.313 ug/l

RT: 10.06 min Scan# 2587

Delta R.T. 0.00 min

Lab File: VN063996.D

Acq: 8 Oct 2020 17:18

Instrument :

MSVOA_N

ClientSampleId :





SA-PZ148-20200930

Tot Ion: 98 Resp: 595074

Ion Ratio Lower Upper

98 100

100 62.8 49.8 74.6



#62

4-Bromofluorobenzene

Concen: 49.482 ug/l

RT: 12.38 min Scan# 3308

Delta R.T. 0.00 min

Lab File: VN063996.D

Acq: 8 Oct 2020 17:18




Tgt Ion: 95 Resp: 224461


Ion Ratio Lower Upper

95 100

174 77.5 0.0 150.6

176 74.1 0.0 145.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.38 min Scan# 2998
Delta R.T. 0.00 min
Lab File: VN063996.D
Acq: 8 Oct 2020 17:18

Instrument :
MSVOA_N
ClientSampleId :
SA-PZ148-20200930

Tot Ion:117 Resp: 476460
Ion Ratio Lower Upper
117 100
82 59.9 47.4 71.2
119 31.2 26.4 39.6




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): V
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.32 min Scan# 3601
Delta R.T. 0.00 min
Lab File: VN063996.D
Acq: 8 Oct 2020 17:18

Tgt Ion:152 Resp: 189891
Ion Ratio Lower Upper
152 100
115 60.7 31.3 93.8
150 156.7 0.0 341.8



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

