

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100820\
 Data File : VN063997.D
 Acq On : 8 Oct 2020 17:42
 Operator : JC/MD
 Sample : L4255-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FC-MW02SR1-20200930

Quant Time: Oct 09 06:34:25 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	249954	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	451422	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	462062	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	180290	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.99	65	208123	50.40	ug/l	0.00
Spiked Amount			Recovery	=	100.80%	
35) Dibromofluoromethane	7.55	113	151736	48.44	ug/l	0.00
Spiked Amount			Recovery	=	96.88%	
50) Toluene-d8	10.06	98	572389	48.20	ug/l	0.00
Spiked Amount			Recovery	=	96.40%	
62) 4-Bromofluorobenzene	12.38	95	217875	48.81	ug/l	0.00
Spiked Amount			Recovery	=	97.62%	


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100820\
Data File : VN063997.D
Acq On : 8 Oct 2020 17:42
Operator : JC/MD
Sample : L4255-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FC-MW02SR1-20200930

Quant Time: Oct 09 06:34:25 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.62 min Scan# 1830

Delta R.T. -0.00 min

Lab File: VN063997.D

Acq: 8 Oct 2020 17:42

Instrument :

MSVOA_N

ClientSampleId :


FC-MW02SR1-20200930

Tot Ion:168 Resp: 249954

Ion Ratio Lower Upper

168 100

99 62.8 51.8 77.6



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN

7.62

120000

100000

80000



60000

40000

20000

0

Time--> 7.50 7.60 7.70



#33

1,2-Dichloroethane-d4

Concen: 50.400 ug/l

RT: 7.99 min Scan# 1943

Delta R.T. 0.00 min

Lab File: VN063997.D


Acq: 8 Oct 2020 17:42

Tgt Ion: 65 Resp: 208123

Ion Ratio Lower Upper

65 100

67 50.4 0.0 99.6



Abundance Ion 64.90 (64.60 to 65.60): VN

Ion 66.90 (66.60 to 67.60): VN

7.99

100000

80000


60000


40000

20000

0

Time--> 7.90 8.00 8.10









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.55 min Scan# 2118
Delta R.T. 0.00 min
Lab File: VN063997.D
Acq: 8 Oct 2020 17:42

Instrument :
MSVOA_N
ClientSampleId :
FC-MW02SR1-20200930

Tot Ion:114 Resp: 451422
Ion Ratio Lower Upper
114 100
63 21.5 0.0 45.0
88 16.3 0.0 34.0




Abundance Ion 113.90 (113.60 to 114.60): V
Ion 63.00 (62.70 to 63.70): VN
Ion 87.90 (87.60 to 88.60): VN






#35
Dibromofluoromethane
Concen: 48.442 ug/l
RT: 7.55 min Scan# 1807
Delta R.T. -0.00 min
Lab File: VN063997.D
Acq: 8 Oct 2020 17:42

Tgt Ion:113 Resp: 151736
Ion Ratio Lower Upper
113 100
111 105.0 81.8 122.8
192 18.8 14.2 21.2



Abundance Ion 112.80 (112.50 to 113.50): V
Ion 110.80 (110.50 to 111.50): V
Ion 191.70 (191.40 to 192.40): V









#50
Toluene-d8
Concen: 48.204 ug/l
RT: 10.06 min Scan# 2588
Delta R.T. 0.00 min
Lab File: VN063997.D
Acq: 8 Oct 2020 17:42




Instrument :
MSVOA_N
ClientSampleId :
FC-MW02SR1-20200930


Tot Ion: 98 Resp: 572389
Ion Ratio Lower Upper
98 100
100 63.3 49.8 74.6



#62
4-Bromofluorobenzene
Concen: 48.811 ug/l
RT: 12.38 min Scan# 3308
Delta R.T. 0.00 min
Lab File: VN063997.D
Acq: 8 Oct 2020 17:42

Tgt Ion: 95 Resp: 217875
Ion Ratio Lower Upper
95 100
174 77.4 0.0 150.6
176 74.9 0.0 145.4





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.38 min Scan# 2998

Delta R.T. 0.00 min

Lab File: VN063997.D

Acq: 8 Oct 2020 17:42

Instrument :

MSVOA_N

ClientSampleId :

FC-MW02SR1-20200930


Tot Ion:117 Resp: 462062

Ion Ratio Lower Upper

117 100

82 60.4 47.4 71.2




119 32.0 26.4 39.6



Abundance Ion 116.90 (116.60 to 117.60): V

Ion 82.00 (81.70 to 82.70): V

Ion 118.90 (118.60 to 119.60): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.32 min Scan# 3601

Delta R.T. 0.00 min

Lab File: VN063997.D

Acq: 8 Oct 2020 17:42


Tgt Ion:152 Resp: 180290

Ion Ratio Lower Upper

152 100

115 62.5 31.3 93.8

150 156.7 0.0 341.8



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

