

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100821\
 Data File : VN068849.D
 Acq On : 8 Oct 2021 15:19
 Operator : JC/MD
 Sample : VN1008WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1008WBL01

Manual Integrations
 APPROVED

MMDadoda
 10/13/2021 12:37:55 PM

Quant Time: Oct 11 04:52:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N092821W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Sep 29 08:56:32 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.662	128	56285	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.971	114	299645	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.747	117	260410	30.00	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.440	65	143358m	33.04	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	110.13%#
60) 4-Bromofluorobenzene	12.736	95	88925	21.86	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	72.87%
63) Toluene-d8	10.443	98	375815	31.93	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	106.43%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100821\
Data File : VN068849.D
Acq On : 8 Oct 2021 15:19
Operator : JC/MD
Sample : VN1008WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1008WBL01

Manual Integrations
APPROVED

MMDadoda
10/13/2021 12:37:55 PM

Quant Time: Oct 11 04:52:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N092821W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Sep 29 08:56:32 2021
Response via : Initial Calibration

