

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100821\
 Data File : VN068854.D
 Acq On : 8 Oct 2021 17:45
 Operator : JC/MD
 Sample : M4105-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 002-35TH-AVE(OCT)

Quant Time: Oct 11 04:54:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N092821W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Sep 29 08:56:32 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.659	128	55961	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.971	114	291669	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.746	117	265606	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.440	65	141020	32.69	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	108.97%
60) 4-Bromofluorobenzene	12.733	95	95177	22.94	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	76.47%
63) Toluene-d8	10.446	98	363108	30.25	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	100.83%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100821\
Data File : VN068854.D
Acq On : 8 Oct 2021 17:45
Operator : JC/MD
Sample : M4105-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
002-35TH-AVE(OCT)

Quant Time: Oct 11 04:54:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N092821W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Sep 29 08:56:32 2021
Response via : Initial Calibration

