

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100824\
 Data File : VN084337.D
 Acq On : 08 Oct 2024 09:43
 Operator : JC\MD
 Sample : VSTDICC150
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC150

Quant Time: Oct 09 02:05:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N100824W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Oct 09 02:04:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.818	128	33784	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.100	114	181536	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.865	117	171821	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.576	65	80830	29.718	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	99.067%
60) 4-Bromofluorobenzene	12.847	95	88349	31.155	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	103.867%
63) Toluene-d8	10.565	98	240116	30.166	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	100.567%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.124	85	301403	146.123	ug/l	96
3) Chloromethane	2.359	50	349225	148.780	ug/l	99
4) Vinyl Chloride	2.512	62	340879	148.089	ug/l	97
5) Bromomethane	2.924	94	208863	136.955	ug/l	99
6) Chloroethane	3.100	64	213547	141.510	ug/l	95
7) Trichlorofluoromethane	3.489	101	546712	145.812	ug/l	99
8) Diethyl Ether	3.959	74	206608	149.881	ug/l	99
9) 1,1,2-Trichlorotrifluo...	4.365	101	314320	149.095	ug/l	98
10) 1,1-Dichloroethene	4.330	96	309616	147.186	ug/l	92
11) Methyl Iodide	4.583	142	422778	152.937	ug/l	99
12) Methyl Acetate	5.018	43	402489	152.836	ug/l	99
13) Acrolein	4.177	56	371221	758.829	ug/l	100
14) Acrylonitrile	5.718	53	902986	763.216	ug/l	99
15) Acetone	4.424	58	277575	821.289	ug/l	95
16) Carbon Disulfide	4.706	76	909776	145.236	ug/l	100
17) Allyl chloride	5.018	41	511759	150.078	ug/l	95
18) Methylene Chloride	5.271	84	349106	147.322	ug/l	98
19) trans-1,2-Dichloroethene	5.783	96	329107	147.827	ug/l	97
20) Diisopropyl ether	6.671	45	1114091	151.956	ug/l	96
21) 1,1-Dichloroethane	6.571	63	620566	147.512	ug/l	97
22) cis-1,2-Dichloroethene	7.482	96	392124	151.363	ug/l	98
23) tert-Butyl Alcohol	5.524	59	333149	734.096	ug/l #	100
24) Methyl tert-Butyl Ether	5.794	73	1068810	152.038	ug/l #	85
25) Chloroform	7.965	83	634973	149.030	ug/l	99
26) Cyclohexane	8.253	56	539286	153.272	ug/l #	97
29) 1,1-Dichloropropene	8.371	75	461334	154.807	ug/l	100
30) 2-Butanone	7.482	43	1249418	787.459	ug/l	99
31) 2,2-Dichloropropane	7.488	77	549477	156.907	ug/l #	80
32) 1,1,1-Trichloroethane	8.165	97	571549	150.045	ug/l	99
33) Carbon Tetrachloride	8.359	117	504637	152.921	ug/l	97
34) Benzene	8.606	78	1425580	152.821	ug/l	96
35) Methacrylonitrile	7.777	41	279509	165.128	ug/l	99
36) 1,2-Dichloroethane	8.671	62	471651	151.686	ug/l	100
37) Trichloroethene	9.353	130	324930	151.266	ug/l	95
38) Methylcyclohexane	9.600	83	538684	164.564	ug/l	99
39) 1,2-Dichloropropane	9.618	63	348207	153.071	ug/l	97
40) Dibromomethane	9.706	93	243118	152.917	ug/l	98
41) Bromodichloromethane	9.888	83	515968	152.872	ug/l	94
42) Vinyl Acetate	6.600	43	4357419	789.646	ug/l	100


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100824\
 Data File : VN084337.D
 Acq On : 08 Oct 2024 09:43
 Operator : JC\MD
 Sample : VSTDICC150
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC150

Quant Time: Oct 09 02:05:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N100824W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Oct 09 02:04:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
43) Ethyl Acetate	7.559	43	484255	154.753	ug/l	100
44) Isopropyl Acetate	8.688	43	808933	157.943	ug/l	99
45) 1,4-Dioxane	9.694	88	147762	3200.393	ug/l	98
46) Methyl methacrylate	9.676	41	391545	158.224	ug/l	100
47) n-amyl Acetate	12.494	43	752246	165.124	ug/l	97
48) t-1,3-Dichloropropene	10.835	75	547316	156.926	ug/l	97
49) cis-1,3-Dichloropropene	10.312	75	578523	158.273	ug/l	98
50) 1,1,2-Trichloroethane	11.018	97	323065	151.179	ug/l	96
51) Ethyl methacrylate	10.870	69	581930	164.863	ug/l	95
52) 1,3-Dichloropropane	11.165	76	568999	152.279	ug/l	99
53) Dibromochloromethane	11.359	129	390970	154.875	ug/l	100
54) 1,2-Dibromoethane	11.470	107	336624	153.033	ug/l	98
55) 2-Chloroethyl vinyl ether	10.159	63	1333727	833.093	ug/l	99
56) Bromoform	12.576	173	268044	160.094	ug/l	99
58) 4-Methyl-2-Pentanone	10.447	43	2491181	778.311	ug/l	99
59) 2-Hexanone	11.194	43	1891152	795.133	ug/l	99
61) Tetrachloroethene	11.106	164	290573	149.507	ug/l	97
62) Toluene	10.629	91	1525116	151.621	ug/l	99
64) Chlorobenzene	11.894	112	946501	151.211	ug/l	100
65) 1,1,1,2-Tetrachloroethane	11.959	131	331211	152.311	ug/l	99
66) Ethyl Benzene	11.964	91	1749567	160.094	ug/l	99
67) m/p-Xylenes	12.070	106	1319448	318.270	ug/l	100
68) o-Xylene	12.394	106	633813	158.611	ug/l	99
69) Styrene	12.412	104	1107376	161.432	ug/l	99
70) Isopropylbenzene	12.694	105	1625283	159.320	ug/l	98
71) 1,1,2,2-Tetrachloroethane	12.935	83	483003	153.053	ug/l	99
72) 1,2,3-Trichloropropane	12.994	75	662180	245.343	ug/l	51
73) Bromobenzene	12.982	156	383306	155.148	ug/l	100
74) n-propylbenzene	13.035	91	1966655	161.778	ug/l	99
75) 2-Chlorotoluene	13.123	91	1199693	158.816	ug/l	99
76) 1,3,5-Trimethylbenzene	13.170	105	1386742	161.447	ug/l	100
77) t-1,4-Dichloro-2-butene	12.735	75	193413	167.468	ug/l	98
78) 4-Chlorotoluene	13.223	91	1227049	159.769	ug/l	100
79) tert-butylbenzene	13.435	119	1182498	161.368	ug/l	99
80) 1,2,4-Trimethylbenzene	13.482	105	1410559	161.849	ug/l	100
81) sec-Butylbenzene	13.617	105	1643555	162.908	ug/l	99
82) p-Isopropyltoluene	13.729	119	1409541	165.992	ug/l	100
83) 1,3-Dichlorobenzene	13.735	146	731253	159.183	ug/l	99
84) 1,4-Dichlorobenzene	13.811	146	723724	158.652	ug/l	99
85) n-Butylbenzene	14.053	91	1274832	173.335	ug/l	98
86) Hexachloroethane	14.335	117	259393	158.553	ug/l	99
87) 1,2-Dichlorobenzene	14.106	146	696639	156.847	ug/l	99
88) 1,2-Dibromo-3-Chloropr...	14.717	75	102940	159.510	ug/l	97
89) 1,2,4-Trichlorobenzene	15.841	180	374861	162.329	ug/l	99
90) Hexachlorobutadiene	15.500	225	157242	147.184	ug/l	97
91) Naphthalene	15.641	128	1334733	170.565	ug/l	99
92) 1,2,3-Trichlorobenzene	15.841	180	374861	162.329	ug/l	99





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Bromochloromethane
Concen: 30.000 ug/l
RT: 7.818 min Scan# 997
Delta R.T. -0.000 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion:128 Resp: 33784
Ion Ratio Lower Upper
128 100
49 185.9 0.0 430.0
130 129.4 0.0 329.8



#2
Dichlorodifluoromethane
Concen: 146.123 ug/l
RT: 2.124 min Scan# 29
Delta R.T. -0.000 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43

Tgt Ion: 85 Resp: 301403
Ion Ratio Lower Upper
85 100
87 33.0 15.4 46.4









#3
Chloromethane
Concen: 148.780 ug/l
RT: 2.359 min Scan# 69
Delta R.T. -0.000 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion: 50 Resp: 349225
Ion Ratio Lower Upper
50 100
52 32.7 25.8 38.6



#4
Vinyl Chloride
Concen: 148.089 ug/l
RT: 2.512 min Scan# 95
Delta R.T. -0.000 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43

Tgt Ion: 62 Resp: 340879
Ion Ratio Lower Upper
62 100
64 32.2 24.3 36.5









#5
Bromomethane
Concen: 136.955 ug/l
RT: 2.924 min Scan# 1
Delta R.T. -0.035 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion: 94 Resp: 208863
Ion Ratio Lower Upper
94 100
96 94.5 76.7 115.1



#6
Chloroethane
Concen: 141.510 ug/l
RT: 3.100 min Scan# 195
Delta R.T. -0.018 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43

Tgt Ion: 64 Resp: 213547
Ion Ratio Lower Upper
64 100
66 31.4 27.7 41.5









#7
Trichlorofluoromethane
Concen: 145.812 ug/l
RT: 3.489 min Scan# 20
Delta R.T. -0.006 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion:101 Resp: 546712
Ion Ratio Lower Upper
101 100
103 60.2 49.0 73.6



#8
Diethyl Ether
Concen: 149.881 ug/l
RT: 3.959 min Scan# 341
Delta R.T. -0.006 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43

Tgt Ion: 74 Resp: 206608
Ion Ratio Lower Upper
74 100
45 95.3 19.2 173.2









#9
1,1,2-Trichlorotrifluoroethane
Concen: 149.095 ug/l
RT: 4.365 min Scan# 410
Delta R.T. -0.000 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion:101 Resp: 314320
Ion Ratio Lower Upper
101 100
85 45.9 0.0 93.6
151 73.4 0.0 151.8



#10
1,1-Dichloroethene
Concen: 147.186 ug/l
RT: 4.330 min Scan# 404
Delta R.T. -0.012 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43

Tgt Ion: 96 Resp: 309616
Ion Ratio Lower Upper
96 100
61 162.6 140.8 211.2
98 62.7 52.2 78.2









#11
Methyl Iodide
Concen: 152.937 ug/l
RT: 4.583 min Scan# 449
Delta R.T. -0.012 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150

Tgt Ion:142 Resp: 422778
Ion Ratio Lower Upper
142 100
127 48.2 23.8 71.4
141 14.5 7.4 22.2



#12
Methyl Acetate
Concen: 152.836 ug/l
RT: 5.018 min Scan# 521
Delta R.T. -0.012 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43

Tgt Ion: 43 Resp: 402489
Ion Ratio Lower Upper
43 100
74 24.8 19.4 29.2









#13
Acrolein
Concen: 758.829 ug/l
RT: 4.177 min Scan# 31
Delta R.T. -0.006 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion: 56 Resp: 371221
Ion Ratio Lower Upper
56 100
55 70.1 56.4 84.6



#14
Acrylonitrile
Concen: 763.216 ug/l
RT: 5.718 min Scan# 640
Delta R.T. -0.000 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43

Tgt Ion: 53 Resp: 902986
Ion Ratio Lower Upper
53 100
52 81.6 40.4 121.2
51 35.6 17.8 53.4









#15
Acetone
Concen: 821.289 ug/l
RT: 4.424 min Scan# 41
Delta R.T. -0.000 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion: 58 Resp: 277575
Ion Ratio Lower Upper
58 100
43 307.5 237.5 356.3



#16
Carbon Disulfide
Concen: 145.236 ug/l
RT: 4.706 min Scan# 468
Delta R.T. -0.000 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43

Tgt Ion: 76 Resp: 909776
Ion Ratio Lower Upper
76 100
78 9.1 7.1 10.7





#17

Allyl chloride

Concen: 150.078 ug/l

RT: 5.018 min Scan# 51

Delta R.T. -0.006 min

Lab File: VN084337.D

Acq: 08 Oct 2024 09:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC150





Tgt Ion: 41 Resp: 511759

Ion Ratio Lower Upper

41 100

39 68.5 31.4 94.3

76 35.0 17.8 53.5



#18

Methylene Chloride

Concen: 147.322 ug/l

RT: 5.271 min Scan# 564

Delta R.T. -0.006 min

Lab File: VN084337.D

Acq: 08 Oct 2024 09:43

Tgt Ion: 84 Resp: 349106




Ion Ratio Lower Upper


84 100

49 120.0 94.6 142.0

51 36.9 27.3 40.9

86 62.3 49.4 74.0









#19
trans-1,2-Dichloroethene
Concen: 147.827 ug/l
RT: 5.783 min Scan# 61
Delta R.T. -0.006 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion: 96 Resp: 329107
Ion Ratio Lower Upper
96 100
61 140.0 115.8 173.6
98 62.5 49.4 74.0



#20
Diisopropyl ether
Concen: 151.956 ug/l
RT: 6.671 min Scan# 802
Delta R.T. -0.000 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43

Tgt Ion: 45 Resp: 1114091
Ion Ratio Lower Upper
45 100
43 50.9 43.7 65.5
87 27.1 23.1 34.7
59 11.8 8.2 12.4









#21
1,1-Dichloroethane
Concen: 147.512 ug/l
RT: 6.571 min Scan# 71
Delta R.T. -0.000 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion: 63 Resp: 620566
Ion Ratio Lower Upper
63 100
98 7.2 4.5 13.4
100 4.2 2.1 6.3



#22
cis-1,2-Dichloroethene
Concen: 151.363 ug/l
RT: 7.482 min Scan# 940
Delta R.T. -0.006 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43

Tgt Ion: 96 Resp: 392124
Ion Ratio Lower Upper
96 100
61 149.9 118.0 177.0
98 63.8 51.8 77.8









#23
tert-Butyl Alcohol
Concen: 734.096 ug/l
RT: 5.524 min Scan# 607
Delta R.T. -0.000 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion: 59 Resp: 333149
Ion Ratio Lower Upper
59 100
52 0.1 0.0 0.0#



#24
Methyl tert-Butyl Ether
Concen: 152.038 ug/l
RT: 5.794 min Scan# 653
Delta R.T. -0.006 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43

Tgt Ion: 73 Resp: 1068810
Ion Ratio Lower Upper
73 100
43 19.8 10.9 16.3#









#25
Chloroform
Concen: 149.030 ug/l
RT: 7.965 min Scan# 1022
Delta R.T. -0.000 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion: 83 Resp: 634973
Ion Ratio Lower Upper
83 100
85 64.4 52.4 78.6



#26
Cyclohexane
Concen: 153.272 ug/l
RT: 8.253 min Scan# 1071
Delta R.T. -0.000 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43

Tgt Ion: 56 Resp: 539286
Ion Ratio Lower Upper
56 100
89 0.0 0.0 0.0
84 86.9 67.0 100.6









#27
 1,2-Dichloroethane-d4
 Concen: 29.718 ug/l
 RT: 8.576 min Scan# 1126
 Delta R.T. -0.006 min
 Lab File: VN084337.D
 Acq: 08 Oct 2024 09:43




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC150


Tgt Ion: 65 Resp: 80830
 Ion Ratio Lower Upper
 65 100
 67 53.4 41.5 62.3



#28
 1,4-Difluorobenzene
 Concen: 30.000 ug/l
 RT: 9.100 min Scan# 1215
 Delta R.T. -0.000 min
 Lab File: VN084337.D
 Acq: 08 Oct 2024 09:43

Tgt Ion: 114 Resp: 181536
 Ion Ratio Lower Upper
 114 100
 63 21.5 17.4 26.0
 88 16.1 13.2 19.8









#29
1,1-Dichloropropene
Concen: 154.807 ug/l
RT: 8.371 min Scan# 110
Delta R.T. -0.000 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion: 75 Resp: 461334
Ion Ratio Lower Upper
75 100
110 34.2 0.0 67.6
77 30.6 0.0 61.4



#30
2-Butanone
Concen: 787.459 ug/l
RT: 7.482 min Scan# 940
Delta R.T. -0.000 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43

Tgt Ion: 43 Resp: 1249418
Ion Ratio Lower Upper
43 100
57 8.3 6.6 9.8
72 27.3 21.3 31.9









#31
2,2-Dichloropropane
Concen: 156.907 ug/l
RT: 7.488 min Scan# 941
Delta R.T. -0.000 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150

Tgt Ion: 77 Resp: 549477
Ion Ratio Lower Upper
77 100
97 22.5 11.5 17.3#



#32
1,1,1-Trichloroethane
Concen: 150.045 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. -0.006 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43

Tgt Ion: 97 Resp: 571549
Ion Ratio Lower Upper
97 100
99 64.5 51.2 76.8
61 43.7 34.1 51.1









#33
Carbon Tetrachloride
Concen: 152.921 ug/l
RT: 8.359 min Scan# 1089
Delta R.T. -0.006 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion:117 Resp: 504637
Ion Ratio Lower Upper
117 100
119 94.9 74.3 111.5
121 30.6 22.4 33.6



#34
Benzene
Concen: 152.821 ug/l
RT: 8.606 min Scan# 1131
Delta R.T. -0.000 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43

Tgt Ion: 78 Resp: 1425580
Ion Ratio Lower Upper
78 100
77 23.2 19.6 29.4
51 17.4 15.8 23.6









#35
Methacrylonitrile
Concen: 165.128 ug/l
RT: 7.777 min Scan# 991
Delta R.T. -0.006 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150

Tgt Ion: 41 Resp: 279509




Ion	Ratio	Lower	Upper
41	100		
39	52.5	26.4	79.0
67	70.6	35.7	107.1
52	27.4	14.5	43.5



#36
1,2-Dichloroethane
Concen: 151.686 ug/l
RT: 8.671 min Scan# 1142
Delta R.T. -0.000 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43

Tgt Ion: 62 Resp: 471651

Ion	Ratio	Lower	Upper
62	100		
98	9.2	7.3	10.9





#37

Trichloroethene

Concen: 151.266 ug/l

RT: 9.353 min Scan# 11

Delta R.T. 0.006 min

Lab File: VN084337.D

Acq: 08 Oct 2024 09:43

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC150

Tgt Ion:130 Resp: 324930

Ion Ratio Lower Upper

130 100

95 101.5 85.8 128.6



#38

Methylcyclohexane

Concen: 164.564 ug/l

RT: 9.600 min Scan# 1300

Delta R.T. -0.000 min

Lab File: VN084337.D

Acq: 08 Oct 2024 09:43




Tgt Ion: 83 Resp: 538684


Ion Ratio Lower Upper

83 100

55 77.3 39.3 117.8

98 47.9 24.0 72.0





#39

1,2-Dichloropropane

Concen: 153.071 ug/l

RT: 9.618 min Scan# 11

Delta R.T. -0.006 min

Lab File: VN084337.D

Acq: 08 Oct 2024 09:43

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC150

Tgt Ion: 63 Resp: 348207

Ion Ratio Lower Upper

63 100

65 31.1 26.3 39.5



#40

Dibromomethane

Concen: 152.917 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. -0.000 min

Lab File: VN084337.D

Acq: 08 Oct 2024 09:43




Tgt Ion: 93 Resp: 243118


Ion Ratio Lower Upper

93 100

95 82.6 0.0 171.4

174 91.2 0.0 183.0









#41
Bromodichloromethane
Concen: 152.872 ug/l
RT: 9.888 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion: 83 Resp: 515968
Ion Ratio Lower Upper
83 100
85 63.9 47.1 70.7
127 8.3 7.1 10.7



#42
Vinyl Acetate
Concen: 789.646 ug/l
RT: 6.600 min Scan# 790
Delta R.T. -0.000 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43

Tgt Ion: 43 Resp: 4357419
Ion Ratio Lower Upper
43 100
86 10.6 8.5 12.7









#43
Ethyl Acetate
Concen: 154.753 ug/l
RT: 7.559 min Scan# 911
Delta R.T. -0.000 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150


Tgt Ion: 43 Resp: 484255
Ion Ratio Lower Upper
43 100
45 13.2 10.6 16.0



#44
Isopropyl Acetate
Concen: 157.943 ug/l
RT: 8.688 min Scan# 1145
Delta R.T. -0.000 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43

Tgt Ion: 43 Resp: 808933
Ion Ratio Lower Upper
43 100
61 26.4 21.4 32.2









#45
1,4-Dioxane
Concen: 3200.393 ug/l
RT: 9.694 min Scan# 11
Delta R.T. -0.006 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion: 88 Resp: 147762
Ion Ratio Lower Upper
88 100
43 29.1 22.4 33.6
58 72.4 59.7 89.5



#46
Methyl methacrylate
Concen: 158.224 ug/l
RT: 9.676 min Scan# 1313
Delta R.T. -0.006 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43

Tgt Ion: 41 Resp: 391545
Ion Ratio Lower Upper
41 100
69 85.4 42.6 128.0
39 55.3 27.5 82.4









#47
n-amyl acetate
Concen: 165.124 ug/l
RT: 12.494 min Scan# 1792
Delta R.T. -0.000 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion: 43 Resp: 752246
Ion Ratio Lower Upper
43 100
55 24.5 20.6 31.0



#48
t-1,3-Dichloropropene
Concen: 156.926 ug/l
RT: 10.835 min Scan# 1510
Delta R.T. -0.000 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43

Tgt Ion: 75 Resp: 547316
Ion Ratio Lower Upper
75 100
77 32.6 24.6 37.0









#49
 cis-1,3-Dichloropropene
 Concen: 158.273 ug/l
 RT: 10.312 min Scan# 1421
 Delta R.T. -0.000 min
 Lab File: VN084337.D
 Acq: 08 Oct 2024 09:43

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC150

Tgt Ion: 75 Resp: 578523




Ion	Ratio	Lower	Upper
75	100		
77	31.8	22.9	34.3
39	46.6	37.4	56.2




#50
 1,1,2-Trichloroethane
 Concen: 151.179 ug/l
 RT: 11.018 min Scan# 1541
 Delta R.T. -0.000 min
 Lab File: VN084337.D
 Acq: 08 Oct 2024 09:43

Tgt Ion: 97 Resp: 323065

Ion	Ratio	Lower	Upper
97	100		
83	86.9	67.8	101.8
85	58.0	44.2	66.2
99	65.6	47.8	71.8









#51
Ethyl methacrylate
Concen: 164.863 ug/l
RT: 10.870 min Scan# 1516
Delta R.T. -0.006 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150


Tgt Ion: 69 Resp: 581930
Ion Ratio Lower Upper
69 100
41 65.7 30.9 92.7
39 35.8 16.6 49.7



#52
1,3-Dichloropropane
Concen: 152.279 ug/l
RT: 11.165 min Scan# 1566
Delta R.T. -0.000 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43

Tgt Ion: 76 Resp: 568999
Ion Ratio Lower Upper
76 100
78 32.6 0.0 64.2









#53
Dibromochloromethane
Concen: 154.875 ug/l
RT: 11.359 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion:129 Resp: 390970
Ion Ratio Lower Upper
129 100
127 77.4 38.5 115.5



#54
1,2-Dibromoethane
Concen: 153.033 ug/l
RT: 11.470 min Scan# 1618
Delta R.T. -0.000 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43

Tgt Ion:107 Resp: 336624
Ion Ratio Lower Upper
107 100
109 94.2 73.8 110.6









#55
2-Chloroethyl vinyl ether
Concen: 833.093 ug/l
RT: 10.159 min Scan# 1395
Delta R.T. -0.000 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150


Tgt Ion: 63 Resp: 1333727
Ion Ratio Lower Upper
63 100
65 32.3 25.6 38.4
106 28.0 21.6 32.4



#56
Bromoform
Concen: 160.094 ug/l
RT: 12.576 min Scan# 1806
Delta R.T. -0.000 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43

Tgt Ion:173 Resp: 268044
Ion Ratio Lower Upper
173 100
175 48.6 39.4 59.2
254 8.1 6.8 10.2





#57

Chlorobenzene-d5

Concen: 30.000 ug/l

RT: 11.865 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN084337.D

Acq: 08 Oct 2024 09:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC150





Tgt Ion:117 Resp: 171821

Ion Ratio Lower Upper

117 100

82 58.8 46.2 69.4

119 32.4 25.4 38.2



#58

4-Methyl-2-Pentanone

Concen: 778.311 ug/l

RT: 10.447 min Scan# 1444

Delta R.T. -0.000 min

Lab File: VN084337.D

Acq: 08 Oct 2024 09:43




Tgt Ion: 43 Resp: 2491181


Ion Ratio Lower Upper

43 100

58 39.6 31.9 47.9

85 18.0 14.6 22.0









#59
2-Hexanone
Concen: 795.133 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. -0.000 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion: 43 Resp: 1891152
Ion Ratio Lower Upper
43 100
58 55.7 44.0 66.0
57 19.3 15.4 23.2



#60
4-Bromofluorobenzene
Concen: 31.155 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43

Tgt Ion: 95 Resp: 88349
Ion Ratio Lower Upper
95 100
174 73.9 58.2 87.2
176 72.8 56.4 84.6









#61
Tetrachloroethene
Concen: 149.507 ug/l
RT: 11.106 min Scan# 11
Delta R.T. 0.006 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion:164 Resp: 290573
Ion Ratio Lower Upper
164 100
166 126.3 102.5 153.7
129 99.1 75.7 113.5
131 93.8 78.2 117.4



#62
Toluene
Concen: 151.621 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43

Tgt Ion: 91 Resp: 1525116
Ion Ratio Lower Upper
91 100
92 58.4 47.2 70.8









#63
Toluene-d8
Concen: 30.166 ug/l
RT: 10.565 min Scan# 1464
Delta R.T. -0.000 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion: 98 Resp: 240116
Ion Ratio Lower Upper
98 100
100 66.6 52.5 78.7



#64
Chlorobenzene
Concen: 151.211 ug/l
RT: 11.894 min Scan# 1690
Delta R.T. 0.006 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43

Tgt Ion: 112 Resp: 946501
Ion Ratio Lower Upper
112 100
114 31.6 25.1 37.7





#65

1,1,1,2-Tetrachloroethane

Concen: 152.311 ug/l

RT: 11.959 min Scan# 1701

Delta R.T. -0.000 min

Lab File: VN084337.D

Acq: 08 Oct 2024 09:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC150





Tgt Ion:131 Resp: 331211

Ion Ratio Lower Upper

131 100

133 94.0 0.0 190.4

119 65.5 0.0 130.0



#66

Ethyl Benzene

Concen: 160.094 ug/l

RT: 11.964 min Scan# 1702

Delta R.T. -0.000 min

Lab File: VN084337.D




Acq: 08 Oct 2024 09:43


Tgt Ion: 91 Resp: 1749567

Ion Ratio Lower Upper

91 100

106 31.1 25.2 37.8









#67
m/p-Xylenes
Concen: 318.270 ug/l
RT: 12.070 min Scan# 1720
Delta R.T. -0.000 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion:106 Resp: 1319448
Ion Ratio Lower Upper
106 100
91 209.0 166.7 250.1



#68
o-Xylene
Concen: 158.611 ug/l
RT: 12.394 min Scan# 1775
Delta R.T. -0.006 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43

Tgt Ion:106 Resp: 633813
Ion Ratio Lower Upper
106 100
91 220.6 110.9 332.9









#69
Styrene
Concen: 161.432 ug/l
RT: 12.412 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion:104 Resp: 1107376
Ion Ratio Lower Upper
104 100
78 52.2 41.6 62.4
103 54.1 44.4 66.6



#70
Isopropylbenzene
Concen: 159.320 ug/l
RT: 12.694 min Scan# 1826
Delta R.T. -0.000 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43

Tgt Ion:105 Resp: 1625283
Ion Ratio Lower Upper
105 100
120 26.0 17.4 32.4









#71
 1,1,2,2-Tetrachloroethane
 Concen: 153.053 ug/l
 RT: 12.935 min Scan# 1867
 Delta R.T. -0.000 min
 Lab File: VN084337.D
 Acq: 08 Oct 2024 09:43




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC150


Tgt Ion: 83 Resp: 483003
 Ion Ratio Lower Upper
 83 100
 131 9.6 5.1 15.3
 85 63.6 32.0 96.2



#72
 1,2,3-Trichloropropane
 Concen: 245.343 ug/l
 RT: 12.994 min Scan# 1877
 Delta R.T. 0.006 min
 Lab File: VN084337.D
 Acq: 08 Oct 2024 09:43

Tgt Ion: 75 Resp: 662180
 Ion Ratio Lower Upper
 75 100
 77 130.8 0.0 414.0









#73
Bromobenzene
Concen: 155.148 ug/l
RT: 12.982 min Scan# 1875
Delta R.T. -0.000 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion:156 Resp: 383306
Ion Ratio Lower Upper
156 100
77 226.0 0.0 451.0
158 96.9 0.0 194.6



#74
n-propylbenzene
Concen: 161.778 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. -0.000 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43

Tgt Ion: 91 Resp: 1966655
Ion Ratio Lower Upper
91 100
120 22.2 0.0 43.2









#75
2-Chlorotoluene
Concen: 158.816 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. -0.000 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150


Tgt Ion: 91 Resp: 1199693
Ion Ratio Lower Upper
91 100
126 31.3 0.0 64.0



#76
1,3,5-Trimethylbenzene
Concen: 161.447 ug/l
RT: 13.170 min Scan# 1907
Delta R.T. -0.000 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43

Tgt Ion: 105 Resp: 1386742
Ion Ratio Lower Upper
105 100
120 47.5 0.0 94.4









#77
t-1,4-Dichloro-2-butene
Concen: 167.468 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. -0.000 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150


Tgt Ion: 75 Resp: 193413
Ion Ratio Lower Upper
75 100
53 92.6 47.3 142.0
89 46.4 22.9 68.8



#78
4-Chlorotoluene
Concen: 159.769 ug/l
RT: 13.223 min Scan# 1916
Delta R.T. 0.006 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43

Tgt Ion: 91 Resp: 1227049
Ion Ratio Lower Upper
91 100
126 31.3 0.0 62.4





#79

tert-butylbenzene

Concen: 161.368 ug/l

RT: 13.435 min Scan# 1952

Delta R.T. -0.000 min

Lab File: VN084337.D

Acq: 08 Oct 2024 09:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC150





Tgt Ion:119 Resp: 1182498

Ion Ratio Lower Upper

119 100

91 67.9 0.0 138.0

134 23.5 0.0 48.8



#80

1,2,4-Trimethylbenzene

Concen: 161.849 ug/l

RT: 13.482 min Scan# 1960

Delta R.T. -0.000 min

Lab File: VN084337.D




Acq: 08 Oct 2024 09:43


Tgt Ion:105 Resp: 1410559

Ion Ratio Lower Upper

105 100

120 44.4 0.0 88.8









#81
 sec-Butylbenzene
 Concen: 162.908 ug/l
 RT: 13.617 min Scan# 1983
 Delta R.T. -0.000 min
 Lab File: VN084337.D
 Acq: 08 Oct 2024 09:43




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC150


Tgt Ion:105 Resp: 1643555
 Ion Ratio Lower Upper
 105 100
 134 19.5 0.0 38.0



#82
 p-Isopropyltoluene
 Concen: 165.992 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. -0.000 min
 Lab File: VN084337.D
 Acq: 08 Oct 2024 09:43

Tgt Ion:119 Resp: 1409541
 Ion Ratio Lower Upper
 119 100
 134 25.3 0.0 50.2
 91 25.0 0.0 50.6









#83
1,3-Dichlorobenzene
Concen: 159.183 ug/l
RT: 13.735 min Scan# 2003
Delta R.T. -0.000 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion:146 Resp: 731253
Ion Ratio Lower Upper
146 100
111 42.6 20.6 61.9
148 63.7 31.6 94.7



#84
1,4-Dichlorobenzene
Concen: 158.652 ug/l
RT: 13.811 min Scan# 2016
Delta R.T. -0.000 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43

Tgt Ion:146 Resp: 723724
Ion Ratio Lower Upper
146 100
111 41.4 20.5 61.5
148 63.3 32.5 97.4









#85
n-Butylbenzene
Concen: 173.335 ug/l
RT: 14.053 min Scan# 2057
Delta R.T. -0.000 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion: 91 Resp: 1274832
Ion Ratio Lower Upper
91 100
92 53.2 0.0 103.4
134 24.8 0.0 48.0



#86
Hexachloroethane
Concen: 158.553 ug/l
RT: 14.335 min Scan# 2105
Delta R.T. -0.000 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43

Tgt Ion: 117 Resp: 259393
Ion Ratio Lower Upper
117 100
201 68.7 33.9 101.6





#87

1,2-Dichlorobenzene

Concen: 156.847 ug/l

RT: 14.106 min Scan# 2066

Delta R.T. -0.000 min

Lab File: VN084337.D

Acq: 08 Oct 2024 09:43

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC150





Tgt Ion:146 Resp: 696639

Ion Ratio Lower Upper

146 100

111 44.0 21.9 65.5

148 63.8 31.1 93.5



#88

1,2-Dibromo-3-Chloropropane

Concen: 159.510 ug/l

RT: 14.717 min Scan# 2170

Delta R.T. -0.006 min

Lab File: VN084337.D

Acq: 08 Oct 2024 09:43




Tgt Ion: 75 Resp: 102940


Ion Ratio Lower Upper

75 100

155 78.1 61.8 92.8

157 101.7 77.4 116.0









#89
1,2,4-Trichlorobenzene
Concen: 162.329 ug/l
RT: 15.841 min Scan# 2361
Delta R.T. -0.000 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150


Tgt Ion:180 Resp: 374861
Ion Ratio Lower Upper
180 100
182 95.8 77.8 116.6
145 35.2 28.9 43.3



#90
Hexachlorobutadiene
Concen: 147.184 ug/l
RT: 15.500 min Scan# 2303
Delta R.T. -0.000 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43

Tgt Ion:225 Resp: 157242
Ion Ratio Lower Upper
225 100
223 64.0 0.0 124.0
227 62.5 0.0 130.8









#91
Naphthalene
Concen: 170.565 ug/l
RT: 15.641 min Scan# 2327
Delta R.T. -0.000 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150

Tgt Ion:128 Resp: 1334733

Ion	Ratio	Lower	Upper
128	100		
127	12.7	10.3	15.5
129	10.8	8.8	13.2



#92
1,2,3-Trichlorobenzene
Concen: 162.329 ug/l
RT: 15.841 min Scan# 2361
Delta R.T. -0.000 min
Lab File: VN084337.D
Acq: 08 Oct 2024 09:43

Tgt Ion:180 Resp: 374861

Ion	Ratio	Lower	Upper
180	100		
182	95.8	0.0	194.4
145	35.2	0.0	72.2

