



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100824\
Data File : VN084336.D
Acq On : 08 Oct 2024 09:08
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N100824W.M
Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
Last Update : Wed Oct 09 02:22:28 2024



AutoFind: Scans 1851, 1852, 1853; Background Corrected with Scan 1844

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.2	8273	PASS
75	95	30	60	50.8	21899	PASS
95	95	100	100	100.0	43144	PASS
96	95	5	9	6.5	2825	PASS
173	174	0.00	2	1.5	464	PASS
174	95	50	100	72.2	31163	PASS
175	174	5	9	7.5	2328	PASS
176	174	95	101	95.4	29744	PASS
177	176	5	9	6.3	1873	PASS