

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100825\
 Data File : VN088035.D
 Acq On : 08 Oct 2025 11:33
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Oct 09 02:40:13 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N100825W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Oct 09 02:36:32 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.800	128	67814	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.082	114	383340	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.847	117	334900	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.565	65	157895	29.901	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	99.667%
60) 4-Bromofluorobenzene	12.829	95	156839	28.408	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	94.700%
63) Toluene-d8	10.547	98	446011	29.891	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	99.633%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100825\
Data File : VN088035.D
Acq On : 08 Oct 2025 11:33
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Oct 09 02:40:13 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N100825W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Oct 09 02:36:32 2025
Response via : Initial Calibration

