

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100918\
 Data File : VN051750.D
 Acq On : 9 Oct 2018 14:31
 Operator : MD\SY
 Sample : J5360-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 2-AVE-22

Quant Time: Oct 10 10:09:58 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N100518W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Oct 05 11:11:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.20	128	113609	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	627580	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	506533	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	8.03	65	224048	28.41	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	94.70%
60) 4-Bromofluorobenzene	12.40	95	184354	23.85	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	79.50%
63) Toluene-d8	10.09	98	722127	32.11	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	107.03%

Target Compounds

					Ovalue
22) cis-1,2-Dichloroethene	6.83	96	124879	17.705 ug/l #	74
37) Trichloroethene	8.84	130	38853	5.538 ug/l	97
61) Tetrachloroethene	10.63	164	144582	24.229 ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100918\
Data File : VN051750.D
Acq On : 9 Oct 2018 14:31
Operator : MD\SY
Sample : J5360-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
2-AVE-22

Quant Time: Oct 10 10:09:58 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N100518W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Oct 05 11:11:26 2018
Response via : Initial Calibration

