

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN100918\
 Data File : VN051754.D
 Acq On : 9 Oct 2018 17:43
 Operator : MD\SY
 Sample : VN1009MBL01
 Misc : 5.00µ/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 13 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1009MBL01

Quant Time: Oct 10 04:15:57 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N100418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 09 06:03:49 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	735418	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1045130	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	907881	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	341543	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	429682	44.23	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.46%	
35) Dibromofluoromethane	7.59	113	388886	48.13	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.26%	
50) Toluene-d8	10.09	98	1476739	48.99	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.98%	
62) 4-Bromofluorobenzene	12.40	95	411930	38.60	ug/l	0.00
Spiked Amount	50.000		Recovery	=	77.20%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100918\
Data File : VN051754.D
Acq On : 9 Oct 2018 17:43
Operator : MD\SY
Sample : VN1009MBL01
Misc : 5.00µ/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 13 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
VN1009MBL01

Quant Time: Oct 10 04:15:57 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N100418W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 09 06:03:49 2018
Response via : Initial Calibration

