

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100918\
 Data File : VN051765.D
 Acq On : 9 Oct 2018 22:15
 Operator : MD\SY
 Sample : PB113645ZHE#05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 24 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB113645ZHE#05

Quant Time: Oct 10 04:43:18 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N100418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 09 06:03:49 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	713636	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1050031	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	925544	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	362323	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	427405	45.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.68%	
35) Dibromofluoromethane	7.60	113	386003	47.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.10%	
50) Toluene-d8	10.09	98	1499778	49.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.04%	
62) 4-Bromofluorobenzene	12.40	95	426725	39.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.60%	
Target Compounds						
18) Methyl Acetate	4.34	43	15847	3.045	ug/l #	83
20) Methylene Chloride	4.56	84	8882	1.145	ug/l	87
43) Isopropyl Acetate	8.17	43	436728	36.594	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100918\
Data File : VN051765.D
Acq On : 9 Oct 2018 22:15
Operator : MD\SY
Sample : PB113645ZHE#05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 24 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB113645ZHE#05

Quant Time: Oct 10 04:43:18 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N100418W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 09 06:03:49 2018
Response via : Initial Calibration

