

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100918\
 Data File : VN051766.D
 Acq On : 9 Oct 2018 22:40
 Operator : MD\SY
 Sample : PB113645ZHE#06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 25 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB113645ZHE#06

Quant Time: Oct 10 04:45:12 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N100418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 09 06:03:49 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	729959	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1058977	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	948806	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	353865	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	433394	44.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.88%	
35) Dibromofluoromethane	7.60	113	396123	48.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.76%	
50) Toluene-d8	10.09	98	1515841	49.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.26%	
62) 4-Bromofluorobenzene	12.40	95	432416	39.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.98%	
Target Compounds						
18) Methyl Acetate	4.34	43	16897	3.175	ug/l	96
20) Methylene Chloride	4.55	84	9202	1.160	ug/l	96
43) Isopropyl Acetate	8.17	43	438779	36.455	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100918\
Data File : VN051766.D
Acq On : 9 Oct 2018 22:40
Operator : MD\SY
Sample : PB113645ZHE#06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 25 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB113645ZHE#06

Quant Time: Oct 10 04:45:12 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N100418W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 09 06:03:49 2018
Response via : Initial Calibration

