

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100919\
 Data File : VN058630.D
 Acq On : 9 Oct 2019 10:13
 Operator : JC/SP
 Sample : K5260-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FA19-GMZ1FLDDS-TB1

Quant Time: Oct 10 01:17:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N100819SIM-VINYL.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 09 04:07:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	1686423	5.00	ug/l	-0.02
4) 1,4-Difluorobenzene	8.56	114	2470958	5.00	ug/l	0.00
8) Chlorobenzene-d5	11.41	117	2206750	5.00	ug/l	0.00
9) 1,4-Dichlorobenzene-d4	13.35	152	802943	5.00	ug/l	0.00

System Monitoring Compounds						
3) 1,2-Dichloroethane-d4	8.00	65	1019181	5.06	ug/l	0.00
Spiked Amount	5.000	Range	81 - 118	Recovery	=	101.20%
5) Dibromofluoromethane	7.57	113	807605	5.19	ug/l	0.00
Spiked Amount	5.000	Range	80 - 119	Recovery	=	103.80%
6) Toluene-d8	10.08	98	1723243	4.90	ug/l	0.00
Spiked Amount	5.000	Range	89 - 112	Recovery	=	98.00%
7) 4-Bromofluorobenzene	12.40	95	772104	4.48	ug/l	0.00
Spiked Amount	5.000	Range	85 - 114	Recovery	=	89.60%

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100919\
Data File : VN058630.D
Acq On : 9 Oct 2019 10:13
Operator : JC/SP
Sample : K5260-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FA19-GMZ1FLDDS-TB1

Quant Time: Oct 10 01:17:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N100819SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Wed Oct 09 04:07:23 2019
Response via : Initial Calibration

