

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100919\
 Data File : VN058639.D
 Acq On : 9 Oct 2019 13:30
 Operator : JC/SP
 Sample : K5260-05MSD
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FA19-MMW0010MSD

Quant Time: Oct 10 01:18:19 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N100819SIM-VINYL.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 09 04:07:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	1542631	5.00	ug/l	0.00
4) 1,4-Difluorobenzene	8.56	114	2285807	5.00	ug/l	0.00
8) Chlorobenzene-d5	11.41	117	2052620	5.00	ug/l	0.00
9) 1,4-Dichlorobenzene-d4	13.35	152	754730	5.00	ug/l	0.00

System Monitoring Compounds

3) 1,2-Dichloroethane-d4	8.00	65	954805	5.18	ug/l	0.00
Spiked Amount	5.000	Range	81 - 118	Recovery	=	103.60%
5) Dibromofluoromethane	7.57	113	752480	5.23	ug/l	0.00
Spiked Amount	5.000	Range	80 - 119	Recovery	=	104.60%
6) Toluene-d8	10.08	98	1602426	4.92	ug/l	0.00
Spiked Amount	5.000	Range	89 - 112	Recovery	=	98.40%
7) 4-Bromofluorobenzene	12.40	95	726089	4.55	ug/l	0.00
Spiked Amount	5.000	Range	85 - 114	Recovery	=	91.00%

Target Compounds

	R.T.	QIon	Response	Conc	Units	Qvalue
2) Vinyl Chloride	2.16	62	789456	5.036	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100919\
Data File : VN058639.D
Acq On : 9 Oct 2019 13:30
Operator : JC/SP
Sample : K5260-05MSD
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FA19-MMW0010MSD

Quant Time: Oct 10 01:18:19 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N100819SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Wed Oct 09 04:07:23 2019
Response via : Initial Calibration

