

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN100919\
Data File : VN058640.D
Acq On : 9 Oct 2019 13:52
Operator : JC/SP
Sample : VSSTDCCC005
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSSTDCCC005

Quant Time: Oct 10 01:18:27 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N100819SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Wed Oct 09 04:07:23 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	5.000	5.000	0.0	51	-0.01
2 C	Vinyl Chloride	5.000	5.130	-2.6#	50	0.00
3 S	1,2-Dichloroethane-d4	5.000	5.274	-5.5	52	0.00
4 I	1,4-Difluorobenzene	5.000	5.000	0.0	51	0.00
5 S	Dibromofluoromethane	5.000	5.375	-7.5	53	0.00
6 S	Toluene-d8	5.000	5.138	-2.8	51	0.00
7 S	4-Bromofluorobenzene	5.000	4.759	4.8	49	0.00
8 I	Chlorobenzene-d5	5.000	5.000	0.0	54	0.00
9 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	49	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1