

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100920\
 Data File : VN064036.D
 Acq On : 9 Oct 2020 14:27
 Operator : JC/MD
 Sample : L4301-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP-BLANK

Quant Time: Oct 10 01:51:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N100620W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Oct 06 16:42:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.15	128	42311	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.55	114	216414	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.38	117	195324	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	7.99	65	114892	30.24	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	100.80%
60) 4-Bromofluorobenzene	12.38	95	80294	24.04	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	80.13%
63) Toluene-d8	10.06	98	263728	29.79	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	99.30%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100920\
Data File : VN064036.D
Acq On : 9 Oct 2020 14:27
Operator : JC/MD
Sample : L4301-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK

Quant Time: Oct 10 01:51:19 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N100620W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Oct 06 16:42:38 2020
Response via : Initial Calibration

