

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100920\
 Data File : VN064041.D
 Acq On : 9 Oct 2020 16:26
 Operator : JC/MD
 Sample : L4327-02 5X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 14B-2

Quant Time: Oct 10 02:01:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N100620W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Oct 06 16:42:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.15	128	43532	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.55	114	214223	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.38	117	200055	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	7.99	65	118663	30.35	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	101.17%
60) 4-Bromofluorobenzene	12.38	95	80085	23.41	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	78.03%
63) Toluene-d8	10.06	98	267438	29.50	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	98.33%

Target Compounds

					Ovalue	
25) Chloroform	7.33	83	395863	71.354	ug/l	99
41) Bromodichloromethane	9.37	83	36647	8.711	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100920\
Data File : VN064041.D
Acq On : 9 Oct 2020 16:26
Operator : JC/MD
Sample : L4327-02 5X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
14B-2

Quant Time: Oct 10 02:01:46 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N100620W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Oct 06 16:42:38 2020
Response via : Initial Calibration

