

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100920\
 Data File : VN064054.D
 Acq On : 9 Oct 2020 23:13
 Operator : JC/MD
 Sample : PB132229TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB132229TB

Quant Time: Oct 10 02:53:59 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	197546	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	365005	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	395137	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	147268	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	167147	51.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.44%	
35) Dibromofluoromethane	7.55	113	123878	48.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.82%	
50) Toluene-d8	10.06	98	482332	50.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.48%	
62) 4-Bromofluorobenzene	12.38	95	183960	50.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.94%	
Target Compounds						
16) Acetone	3.78	43	25517	25.399	ug/l	97
20) Methylene Chloride	4.51	84	26136	11.434	ug/l	98
43) Isopropyl Acetate	8.13	43	75046	12.384	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100920\
Data File : VN064054.D
Acq On : 9 Oct 2020 23:13
Operator : JC/MD
Sample : PB132229TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB132229TB

Quant Time: Oct 10 02:53:59 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

