

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100920\
 Data File : VN064058.D
 Acq On : 10 Oct 2020 00:49
 Operator : JC/MD
 Sample : PB132229ZHE#12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB132229ZHE#12

Quant Time: Oct 10 03:01:33 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	207750	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	416989	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	433845	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	160680	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	186832	54.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.88%	
35) Dibromofluoromethane	7.55	113	135573	46.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.72%	
50) Toluene-d8	10.06	98	535886	48.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.72%	
62) 4-Bromofluorobenzene	12.38	95	199790	48.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.92%	
Target Compounds						
16) Acetone	3.78	43	26855	25.418	ug/l	92
20) Methylene Chloride	4.51	84	27361	11.382	ug/l	98
43) Isopropyl Acetate	8.13	43	75355	10.885	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100920\
Data File : VN064058.D
Acq On : 10 Oct 2020 00:49
Operator : JC/MD
Sample : PB132229ZHE#12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB132229ZHE#12

Quant Time: Oct 10 03:01:33 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

