

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100920\
 Data File : VN064062.D
 Acq On : 10 Oct 2020 2:25
 Operator : JC/MD
 Sample : PB132229ZHE#16
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB132229ZHE#16

Manual Integrations
 APPROVED

MMDadoda
 10/12/2020 2:10:17 PM

Quant Time: Oct 10 03:10:11 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	201906	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	403545	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	420176	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	151589	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	178505	53.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.02%	
35) Dibromofluoromethane	7.55	113	132571	47.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.70%	
50) Toluene-d8	10.06	98	518245	48.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.64%	
62) 4-Bromofluorobenzene	12.38	95	194063	48.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.26%	
Target Compounds						
16) Acetone	3.78	43	13600	13.245	ug/l	92
18) Methyl Acetate	4.28	43	4152	1.557	ug/l #	56
20) Methylene Chloride	4.51	84	6708m	2.945	ug/l	
43) Isopropyl Acetate	8.13	43	86363	12.891	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100920\
Data File : VN064062.D
Acq On : 10 Oct 2020 2:25
Operator : JC/MD
Sample : PB132229ZHE#16
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB132229ZHE#16

Manual Integrations
APPROVED

MMDadoda
10/12/2020 2:10:17 PM

Quant Time: Oct 10 03:10:11 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

