

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100920\
 Data File : VN064068.D
 Acq On : 10 Oct 2020 4:49
 Operator : JC/MD
 Sample : L4319-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 11995

Quant Time: Oct 10 05:55:44 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	199074	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	400415	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	414025	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	156909	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	181826	55.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.58%	
35) Dibromofluoromethane	7.55	113	133123	47.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.82%	
50) Toluene-d8	10.06	98	518948	49.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.54%	
62) 4-Bromofluorobenzene	12.38	95	193568	48.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.78%	
Target Compounds						
16) Acetone	3.78	43	41322	40.815	ug/l	95
20) Methylene Chloride	4.51	84	10869	4.776	ug/l	92
43) Isopropyl Acetate	8.13	43	74054	11.140	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100920\
Data File : VN064068.D
Acq On : 10 Oct 2020 4:49
Operator : JC/MD
Sample : L4319-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
11995

Quant Time: Oct 10 05:55:44 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

