

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100920\
 Data File : VN064069.D
 Acq On : 10 Oct 2020 5:13
 Operator : JC/MD
 Sample : L4319-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 12018

Quant Time: Oct 10 05:57:50 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	201505	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	407503	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	420559	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	160647	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	182438	54.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.60%	
35) Dibromofluoromethane	7.55	113	133061	47.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.12%	
50) Toluene-d8	10.06	98	513270	47.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.76%	
62) 4-Bromofluorobenzene	12.38	95	193979	48.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.28%	
Target Compounds						
16) Acetone	3.78	43	45215	44.122	ug/l	95
18) Methyl Acetate	4.29	43	4259	1.601	ug/l #	84
20) Methylene Chloride	4.50	84	10808	4.694	ug/l #	85
43) Isopropyl Acetate	8.13	43	73302	10.835	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100920\
Data File : VN064069.D
Acq On : 10 Oct 2020 5:13
Operator : JC/MD
Sample : L4319-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
12018

Quant Time: Oct 10 05:57:50 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

