

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100923\
 Data File : VN079468.D
 Acq On : 09 Oct 2023 13:35
 Operator : JC\MD
 Sample : 04731-09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DTP-3

Quant Time: Oct 09 23:14:20 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	269765	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	547369	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	557190	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	170901	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	228540	52.249	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.500%
35) Dibromofluoromethane	8.171	113	200520	50.976	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.960%
50) Toluene-d8	10.571	98	680861	47.837	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.680%
62) 4-Bromofluorobenzene	12.853	95	226859	49.667	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.340%
Target Compounds						
						Qvalue
16) Acetone	4.436	43	23528	13.562	ug/l	99
20) Methylene Chloride	5.283	84	15884	3.622	ug/l #	81
43) Isopropyl Acetate	8.694	43	212138	20.612	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100923\
Data File : VN079468.D
Acq On : 09 Oct 2023 13:35
Operator : JC\MD
Sample : 04731-09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DTP-3

Quant Time: Oct 09 23:14:20 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration

