

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100923\
 Data File : VN079469.D
 Acq On : 09 Oct 2023 13:59
 Operator : JC\MD
 Sample : 04765-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT-1

Quant Time: Oct 09 23:14:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	270274	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	545115	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	553813	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	171782	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	228131	52.058	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 104.120%	
35) Dibromofluoromethane	8.165	113	197856	50.507	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.020%	
50) Toluene-d8	10.571	98	676216	47.707	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 95.420%	
62) 4-Bromofluorobenzene	12.853	95	227443	50.001	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 100.000%	
Target Compounds						
						Qvalue
16) Acetone	4.441	43	26347	15.158	ug/l	91
20) Methylene Chloride	5.271	84	16713	3.835	ug/l #	92
43) Isopropyl Acetate	8.694	43	209126	20.404	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100923\
Data File : VN079469.D
Acq On : 09 Oct 2023 13:59
Operator : JC\MD
Sample : 04765-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT-1

Quant Time: Oct 09 23:14:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration

