

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100923\
 Data File : VN079471.D
 Acq On : 09 Oct 2023 14:47
 Operator : JC\MD
 Sample : 04739-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW12

Quant Time: Oct 09 23:15:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	252758	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	505037	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	469091	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	146282	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	215370	52.551	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.100%
35) Dibromofluoromethane	8.171	113	187815	51.749	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.500%
50) Toluene-d8	10.571	98	595203	45.324	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.640%
62) 4-Bromofluorobenzene	12.853	95	195082	46.290	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.580%
Target Compounds						
					Qvalue	
47) Bromodichloromethane	9.888	83	5417	0.954	ug/l #	85
60) Dibromochloromethane	11.365	129	17256	4.491	ug/l	100
65) Chlorobenzene	11.894	112	19749	1.890	ug/l	93
71) Bromoform	12.582	173	20647	7.375	ug/l #	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100923\
Data File : VN079471.D
Acq On : 09 Oct 2023 14:47
Operator : JC\MD
Sample : 04739-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW12

Quant Time: Oct 09 23:15:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration

